

Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137273

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/23/2025	08:30
2	CAL2	1	Water	NA	NA	20.2	7.00	09/23/2025	08:31
3	CAL3	1	Water	NA	NA	20.3	10.02	09/23/2025	08:33
4	ICV	1	Water	NA	NA	20.2	6.98	09/23/2025	08:35
5	CCV1	1	Water	NA	NA	20.3	2.01	09/23/2025	08:37
6	Q3160-01	1	Solid	20.02	20	21.2	5.91	09/23/2025	08:45
7	Q3160-01DUP	1	Solid	20.03	20	21.3	5.92	09/23/2025	08:47
8	CCV2	1	Water	NA	NA	20.3	12.02	09/23/2025	08:50

WORKLIST(Hardcopy Internal Chain)

VB 13-12-13

WorkList Name : ph s q3160 WorkList ID : 192010 Department : Wet-Chemistry Date : 09-23-2025 08:13:21

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3160-01	☒ AUD-25-0150	Solid	pH	Cool 4 deg C	PSEG03	D31	09/22/2025	9045D

Date/Time 09/23/25 08:25
 Raw Sample Received by: JB WWCJ
 Raw Sample Relinquished by: CPSON

Date/Time 09/22/25 10:00
 Raw Sample Received by: CPSON
 Raw Sample Relinquished by: JB WWCJ