

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137290

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	09/24/2025	12:25
2	Q3138-05		1	14	100.00	>212.0	>212.0	09/24/2025	12:55
3	Q3138-05DUP		1	14	100.00	>212.0	>212.0	09/24/2025	13:25
4	Q3138-06		1	15	100.00	>212.0	>212.0	09/24/2025	13:55
5	Q3139-01		1	13	100.00	>212.0	>212.0	09/24/2025	14:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137290

WorkList Name : fp-9-23 WorkList ID : 192013 Department : Wet-Chemistry Date : 09-23-2025 08:19:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3138-05	MOO-25-0264	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/18/2025	1010B
Q3138-06	MOO-25-0242	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/18/2025	1010B
Q3139-01	LAW-25-0140	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/18/2025	1010B

Date/Time 09/24/2025 12.15
 Raw Sample Received by: RM (w/c)
 Raw Sample Relinquished by: [Signature]

Date/Time 09/24/2025 15:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: RM (w/c)