

LB137291


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

26431



A= 0.0000e+000

B= 3.6462e-004

C= -2.2156e-002

Rho= 0.9977310

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.044	-0.044	-59	0.000	-59				
0.05	0.050									
0.20	0.200	0.218	0.018	660	0.0 %	660				
2.50	2.500	2.336	-0.164	6468	0.0 %	6468				
5.00	5.000	5.168	0.168	14234	0.0 %	14234				
7.50	7.500	7.906	0.406	21744	0.0 %	21744				
10.0	10.000	9.615	-0.385	26431	0.0 %	26431				

6/61

10.000
-4.4299

LB137291
INSTRUMENT ID: CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-59	-		0 SFAM01.1	PPB	9/24/2025 13:42	Std
	0.2 S02	660	-		0.2 SFAM01.1	PPB	9/24/2025 13:44	Std
	2.5 S2.5	6468	-		2.5 SFAM01.1	PPB	9/24/2025 13:49	Std
	5 S5.0	14234	-		5 SFAM01.1	PPB	9/24/2025 13:54	Std
	7.5 S7.5	21744	-		7.5 SFAM01.1	PPB	9/24/2025 13:59	Std
	10 S10	26431	-		10 SFAM01.1	PPB	9/24/2025 14:04	Std
ICV44	ICV44	10593	3.8402	-	SFAM01.1	PPB	9/24/2025 14:10	SMPL
ICB44	ICB44	-190	-0.0914	-	SFAM01.1	PPB	9/24/2025 14:12	SMPL
CCV33	CCV33	14026	5.092	-	SFAM01.1	PPB	9/24/2025 14:15	SMPL
CCB33	CCB33	-222	-0.1031	-	SFAM01.1	PPB	9/24/2025 14:17	SMPL
PB169817BL	PBW	7	-0.0196	-	SFAM01.1	PPB	9/24/2025 14:22	SMPL
Q3084-01	PMW-9S(20250909)	366	0.1113	-	SFAM01.1	PPB	9/24/2025 14:24	SMPL
Q3084-02	PMW-9D(20250909)	1216	0.4212	-	SFAM01.1	PPB	9/24/2025 14:26	SMPL
Q3084-03	PMW-8D(20250909)	371	0.1131	-	SFAM01.1	PPB	9/24/2025 14:29	SMPL
Q3084-04	PMW-8S(20250909)	216	0.0566	-	SFAM01.1	PPB	9/24/2025 14:31	SMPL
Q3084-05	FB(20250909)	-127	-0.0685	-	SFAM01.1	PPB	9/24/2025 14:33	SMPL
Q3084-06	PMW-6(20250910)	631	0.2079	-	SFAM01.1	PPB	9/24/2025 14:35	SMPL
Q3084-07	PMW-3(20250910)	190	0.0471	-	SFAM01.1	PPB	9/24/2025 14:38	SMPL
Q3084-08	PMW-7S(20250910)	-7	-0.0247	-	SFAM01.1	PPB	9/24/2025 14:40	SMPL
Q3084-09	PMW-7D(20250910)	68	0.0026	-	SFAM01.1	PPB	9/24/2025 14:42	SMPL
Q3084-10	DUP(20250910)	40	-0.0076	-	SFAM01.1	PPB	9/24/2025 14:45	SMPL
Q3084-11	PMW-10(20250910)	426	0.1332	-	SFAM01.1	PPB	9/24/2025 14:47	SMPL
Q3084-12	FB(20250910)	-95	-0.0568	-	SFAM01.1	PPB	9/24/2025 14:49	SMPL
Q3084-13	PMW-4(20250911)	116	0.0201	-	SFAM01.1	PPB	9/24/2025 14:52	SMPL
Q3084-14	MW-1(MCUA)(20250911)	-31	-0.0335	-	SFAM01.1	PPB	9/24/2025 14:54	SMPL
Q3084-15	MW-2(MCUA)(20250911)	-9	-0.0254	-	SFAM01.1	PPB	9/24/2025 14:56	SMPL
Q3084-16	PMW-1(20250911)	129	0.0249	-	SFAM01.1	PPB	9/24/2025 14:59	SMPL
Q3084-17	PMW-5(20250911)	116	0.0201	-	SFAM01.1	PPB	9/24/2025 15:01	SMPL
Q3084-18	PMW-2(20250911)	87	0.0096	-	SFAM01.1	PPB	9/24/2025 15:03	SMPL
Q3084-19	Q3084-18S	2944	1.0513	-	SFAM01.1	PPB	9/24/2025 15:06	SMPL
CCV34	CCV34	14065	5.1062	-	SFAM01.1	PPB	9/24/2025 15:08	SMPL
CCB34	CCB34	-222	-0.1031	-	SFAM01.1	PPB	9/24/2025 15:10	SMPL
Q3084-20	Q3084-18D	183	0.0446	-	SFAM01.1	PPB	9/24/2025 15:12	SMPL
Q3084-21	FB(20250911)	-38	-0.036	-	SFAM01.1	PPB	9/24/2025 15:15	SMPL
Q3084-22	EB(20250911)	25	-0.013	-	SFAM01.1	PPB	9/24/2025 15:17	SMPL
CCV35	CCV35	14107	5.1215	-	SFAM01.1	PPB	9/24/2025 15:21	SMPL
CCB35	CCB35	-159	-0.0801	-	SFAM01.1	PPB	9/24/2025 15:23	SMPL