

Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137300

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/25/2025	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	09/25/2025	08:01
3	CAL3	1	Water	NA	NA	20.3	10.02	09/25/2025	08:05
4	ICV	1	Water	NA	NA	20.2	7.01	09/25/2025	08:11
5	CCV1	1	Water	NA	NA	20.3	2.01	09/25/2025	08:15
6	Q3168-05	1	Water	NA	NA	20.2	5.75	09/25/2025	08:30
7	Q3168-05DUP	1	Water	NA	NA	20.3	5.76	09/25/2025	08:32
8	Q3168-06	1	Water	NA	NA	20.2	6.30	09/25/2025	08:40
9	CCV2	1	Water	NA	NA	20.3	12.02	09/25/2025	08:44

WORKLIST(Hardcopy Internal Chain)

W131300

WorkList Name : ph w q3151-2 WorkList ID : 192058 Department : Wet-Chemistry Date : 09-25-2025 07:43:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3168-05	TOTES COMP#1	Water	pH	Cool 4 deg C	PSEG03	J11	09/23/2025	9040C
Q3168-06	TOTES COMP#2	Water	pH	Cool 4 deg C	PSEG03	J11	09/23/2025	9040C

Date/Time 09/25/25 07:50
Raw Sample Received by: SA WWC
Raw Sample Relinquished by: CP SR

Date/Time 09/25/25 11:00
Raw Sample Received by: CP SR
Raw Sample Relinquished by: SA WWC