

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137312

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	09/25/2025	14:30
2	Q3168-05		1	13	100.00	>212.0	>212.0	09/25/2025	15:00
3	Q3168-05DUP		1	13	100.00	>212.0	>212.0	09/25/2025	15:30
4	Q3168-06		1	15	100.00	>212.0	>212.0	09/25/2025	16:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137312

WorkList Name : fp-9-24 WorkList ID : 192057 Department : Wet-Chemistry Date : 09-24-2025 16:21:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3168-05	TOTES COMP#1	Water	Flash Point	Cool 4 deg C	PSEG03	J11	09/23/2025	1010B
Q3168-06	TOTES COMP#2	Water	Flash Point	Cool 4 deg C	PSEG03	J11	09/23/2025	1010B

Date/Time 09/25/2025 14:00
 Raw Sample Received by: R14 CWSJ
 Raw Sample Relinquished by: JAWOR

Date/Time 09/25/2025 16:40
 Raw Sample Received by: JAWOR
 Raw Sample Relinquished by: R14 CWSJ