


LB137331
INSTRUMENT ID: CV₁

7471B

Linear

μ Abs.:

46549



A= 0.0000e+000

B= 2.1402e-004

C= 4.8043e-002

Rho= 0.9999034

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.040	0.040	-39	0.000	-39					13
0.2	0.200	0.226	0.026	833	0.0 %	833					0
2.5	2.500	2.455	-0.045	11247	0.0 %	11247					-2
5.0	5.000	4.910	-0.090	22720	0.0 %	22720					1
7.5	7.500	7.558	0.058	35091	0.0 %	35091					0
10.0	10.000	10.010	0.010	46549	0.0 %	46549					

LB137331 INSTRUMENT ID:CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Sep 2025 11:38:59

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50.0	-	-39	PPB	Std	0.0000	7471B	26 Sep 2025 11:56:01
S	0.2 - 1	50.2	-	833	PPB	Std	0.2000	7471B	26 Sep 2025 11:58:18
S	2.5 - 1	52.5	-	11247	PPB	Std	2.5000	7471B	26 Sep 2025 12:00:35
S	5.0 - 1	55.0	-	22720	PPB	Std	5.0000	7471B	26 Sep 2025 12:02:52
S	7.5 - 1	57.5	-	35091	PPB	Std	7.5000	7471B	26 Sep 2025 12:05:22
S	10.0 - 1	510.0	-	46549	PPB	Std	10.0000	7471B	26 Sep 2025 12:08:04
U	ICV48 - 1	ICV48	3.9046	18020	PPB	SMPL	-7471B		26 Sep 2025 12:11:27
U	ICB48 - 1	ICB48	-0.0211	-323	PPB	SMPL	-7471B		26 Sep 2025 12:13:45
U	CCV45 - 1	CCV45	4.9000	22671	PPB	SMPL	-7471B		26 Sep 2025 12:16:23
U	CCB45 - 1	CCB45	-0.0010	-229	PPB	SMPL	-7471B		26 Sep 2025 12:18:38
U	CRA - 1	CRA	0.2111	762	PPB	SMPL	-7471B		26 Sep 2025 12:21:17
U	HighStd - 1	HighStd	9.9412	46226	PPB	SMPL	-7471B		26 Sep 2025 12:23:32
U	ChkStd - 1	ChkStd	6.9636	32313	PPB	SMPL	-7471B		26 Sep 2025 12:25:48
U	PB169861BL - 1	PBS	-0.0119	-280	PPB	SMPL	-7471B		26 Sep 2025 12:28:34
U	PB169861BS - 1	LCSS	3.5209	16227	PPB	SMPL	-7471B		26 Sep 2025 12:31:17
U	Q3200-01 - 1	ETGI-367	0.1420	439	PPB	SMPL	-7471B		26 Sep 2025 12:33:33
U	Q3206-01 - 1	TR-06-092525	0.3372	1351	PPB	SMPL	-7471B		26 Sep 2025 12:36:10
U	Q3208-01 - 1	SU-ESM-COMP-01	0.9107	4031	PPB	SMPL	-7471B		26 Sep 2025 12:38:27
U	Q3208-07 - 1	SU-ESM-COMP-02	2.2584	10328	PPB	SMPL	-7471B		26 Sep 2025 12:40:45
U	Q3209-01 - 1	PL-02-092525	0.2077	746	PPB	SMPL	-7471B		26 Sep 2025 12:43:02
U	CCV46 - 1	CCV46	4.9037	22688	PPB	SMPL	-7471B		26 Sep 2025 12:45:32
U	CCB46 - 1	CCB46	0.0076	-189	PPB	SMPL	-7471B		26 Sep 2025 12:47:47
U	Q3210-01 - 1	SG-1	0.0861	178	PPB	SMPL	-7471B		26 Sep 2025 12:50:27
U	Q3212-01 - 1	EO-03-092525	2.0898	9540	PPB	SMPL	-7471B		26 Sep 2025 12:52:42
U	Q3212-01DUP - 1	EO-03-092525DUP	2.5508	11694	PPB	SMPL	-7471B		26 Sep 2025 12:57:54
U	Q3212-01MS - 1	EO-03-092525MS	6.6151	30685	PPB	SMPL	-7471B		26 Sep 2025 13:00:12
U	Q3212-01MSD - 1	EO-03-092525MSD	5.5361	25643	PPB	SMPL	-7471B		26 Sep 2025 13:02:42
U	Q3212-01LX5 - 1		0.4523	1889	PPB	SMPL	-7471B		26 Sep 2025 13:05:23
U	Q3212-01A - 1		3.8753	17883	PPB	SMPL	-7471B		26 Sep 2025 13:08:04
U	CCV47 - 1	CCV47	4.8803	22579	PPB	SMPL	-7471B		26 Sep 2025 13:10:22
U	CCB47 - 1	CCB47	-0.0042	-244	PPB	SMPL	-7471B		26 Sep 2025 13:12:59