

Analysis Method: 9034

Parameter: Reactive Sulfide

Run Number: LB137334

ANALYST: Iwona

SUPERVISOR REVIEW BY: JIGNESH

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W3105
IODINE SOLUTION .025N 1L	W3213
Starch Solution, 4L	W3149

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB169869BL		1	5.00	50	2.00	0.00	1.96	1.96	0.04	0.00	0.00	09/26/2025	15:30
2	Q3199-01		1	5.07	50	2.00	0.00	1.90	1.90	0.10	0.06	4.73	09/26/2025	15:33
3	Q3199-01DUP		1	5.07	50	2.00	0.00	1.90	1.90	0.10	0.06	4.73	09/26/2025	15:36
4	Q3199-02		1	5.01	50	2.00	0.00	1.92	1.92	0.08	0.04	3.19	09/26/2025	15:40
5	Q3199-03		1	5.03	50	2.00	0.00	1.90	1.90	0.10	0.06	4.77	09/26/2025	15:43
6	Q3199-04		1	5.04	50	2.00	0.00	1.94	1.94	0.06	0.02	1.59	09/26/2025	15:47
7	Q3199-05		1	5.01	50	2.00	0.00	1.88	1.88	0.12	0.08	6.39	09/26/2025	15:50
8	Q3199-06		1	5.07	50	2.00	0.00	1.92	1.92	0.08	0.04	3.16	09/26/2025	15:53
9	Q3199-07		1	5.06	50	2.00	0.00	1.92	1.92	0.08	0.04	3.16	09/26/2025	15:57

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume