


LB137343
INSTRUMENT ID: CU₁

7470A

μ Abs.:

35801



Linear

A= 0.0000e+000

B= 2.7118e-004

C= 5.0059e-003

Rho= 0.9986687

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.003	0.003	-7	0.000	-7	0				-
0.2	0.200	0.171	-0.029	612	0.0 %	612	0				
2.5	2.500	2.418	-0.082	8900	0.0 %	8900	0				
5.0	5.000	5.039	0.039	18563	0.0 %	18563					
7.5	7.500	7.855	0.355	28949	0.0 %	28949					
10.0	10.000	9.713	-0.287	35801	0.0 %	35801					

LB137343 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Sep 2025 10:12:07

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std CondMethod	Date
S	0.0 - 1	50	-	-7	PPB	Std	0.00007470A	29 Sep 2025 10:29:34
S	0.2 - 1	50.2	-	612	PPB	Std	0.20007470A	29 Sep 2025 10:31:52
S	2.5 - 1	52.5	-	8900	PPB	Std	2.50007470A	29 Sep 2025 10:34:09
S	5.0 - 1	53.0	-	18563	PPB	Std	5.00007470A	29 Sep 2025 10:36:26
S	7.5 - 1	57.5	-	28949	PPB	Std	7.50007470A	29 Sep 2025 10:38:45
S	10.0 - 1	510.0	-	35801	PPB	Std	10.00007470A	29 Sep 2025 10:41:04
U	ICV50 - 1	ICV50	3.7141	13678	PPB	SMPL	-7470A	29 Sep 2025 10:44:42
U	ICB50 - 1	ICB50	-0.0541	-218	PPB	SMPL	-7470A	29 Sep 2025 10:46:57
U	CCV51 - 1	CCV51	5.0215	18499	PPB	SMPL	-7470A	29 Sep 2025 10:49:15
U	CCB51 - 1	CCB51	-0.0704	-278	PPB	SMPL	-7470A	29 Sep 2025 10:51:30
U	CRA - 1	CRA	0.1739	623	PPB	SMPL	-7470A	29 Sep 2025 10:56:27
U	HighStd - 1	HighStd	9.6602	35605	PPB	SMPL	-7470A	29 Sep 2025 10:58:41
U	ChkStd - 1	ChkStd	6.9178	25492	PPB	SMPL	-7470A	29 Sep 2025 11:00:57
U	PB169867BL - 1	PBW	-0.0864	-337	PPB	SMPL	-7470A	29 Sep 2025 11:03:22
U	PB169867BS - 1	LCSW	4.0941	15079	PPB	SMPL	-7470A	29 Sep 2025 11:08:34
U	Q3181-03 - 1	WC-A3-01-C	-0.0644	-256	PPB	SMPL	-7470A	29 Sep 2025 11:10:50
U	Q3181-07 - 1	WC-A3-02-C	-0.0026	-28	PPB	SMPL	-7470A	29 Sep 2025 11:13:09
U	Q3181-11 - 1	WC-A3-03-C	-0.0023	-27	PPB	SMPL	-7470A	29 Sep 2025 11:15:26
U	Q3183-10 - 1	WC-1	-0.0004	-20	PPB	SMPL	-7470A	29 Sep 2025 11:17:43
U	Q3183-11 - 1	WC-4	0.0215	61	PPB	SMPL	-7470A	29 Sep 2025 11:19:59
U	CCV52 - 1	CCV52	5.0906	18754	PPB	SMPL	-7470A	29 Sep 2025 11:22:15
U	CCB52 - 1	CCB52	-0.0384	-160	PPB	SMPL	-7470A	29 Sep 2025 11:24:30
U	Q3183-12 - 1	WC-3	0.0072	8	PPB	SMPL	-7470A	29 Sep 2025 11:26:48
U	Q3188-05 - 1	4135	0.0028	-8	PPB	SMPL	-7470A	29 Sep 2025 11:29:05
U	Q3191-01 - 1	459	0.0449	147	PPB	SMPL	-7470A	29 Sep 2025 11:31:36
U	Q3196-01 - 1	FMC-9-24-25-1	-0.0045	-35	PPB	SMPL	-7470A	29 Sep 2025 11:33:52
U	Q3196-02 - 1	FMC-9-24-25-2	0.0351	111	PPB	SMPL	-7470A	29 Sep 2025 11:36:44
U	Q3196-03 - 1	FMC-9-24-25-3	-0.0123	-64	PPB	SMPL	-7470A	29 Sep 2025 11:39:44
U	Q3196-04 - 1	FMC-9-24-25-4	0.0009	-15	PPB	SMPL	-7470A	29 Sep 2025 11:42:00
U	Q3196-05 - 1	FMC-9-24-25-5	0.0018	-12	PPB	SMPL	-7470A	29 Sep 2025 11:44:17
U	Q3208-06 - 1	SU-ESM-COMP-01	0.0161	41	PPB	SMPL	-7470A	29 Sep 2025 11:46:35
U	Q3208-12 - 1	SU-ESM-COMP-02	0.0074	9	PPB	SMPL	-7470A	29 Sep 2025 11:48:51
U	CCV53 - 1	CCV53	4.8439	17844	PPB	SMPL	-7470A	29 Sep 2025 11:51:07
U	CCB53 - 1	CCB53	-0.0330	-140	PPB	SMPL	-7470A	29 Sep 2025 11:53:22
U	Q3208-12DUP - 1	SU-ESM-COMP-02DUP	0.0584	197	PPB	SMPL	-7470A	29 Sep 2025 11:55:41
U	Q3208-12MS - 1	SU-ESM-COMP-02MS	4.2080	15499	PPB	SMPL	-7470A	29 Sep 2025 11:57:57
U	Q3208-12MSD - 1	SU-ESM-COMP-02MSD	4.3579	16052	PPB	SMPL	-7470A	29 Sep 2025 12:00:12
U	PB169868BL - 1	PBW	-0.0206	-94	PPB	SMPL	-7470A	29 Sep 2025 12:02:31
U	PB169868BS - 1	LCSW	4.6473	17119	PPB	SMPL	-7470A	29 Sep 2025 12:04:51
U	Q3196-06 - 1	TOTE-FMC-9-24-25-1	-0.0359	-151	PPB	SMPL	-7470A	29 Sep 2025 12:07:08
U	Q3196-07 - 1	TOTE-FMC-9-24-25-2	0.0262	78	PPB	SMPL	-7470A	29 Sep 2025 12:09:28
U	Q3196-09 - 1	FMC-9-24-25-DRUM-1	0.0050	0	PPB	SMPL	-7470A	29 Sep 2025 12:11:45
U	Q3199-09 - 1	ELZ-25-000107	0.0091	15	PPB	SMPL	-7470A	29 Sep 2025 12:14:02
U	Q3199-09DUP - 1	ELZ-25-000107DUP	0.0131	30	PPB	SMPL	-7470A	29 Sep 2025 12:16:17
U	CCV54 - 1	CCV54	4.6500	17129	PPB	SMPL	-7470A	29 Sep 2025 12:18:33
U	CCB54 - 1	CCB54	-0.0110	-59	PPB	SMPL	-7470A	29 Sep 2025 12:20:49
U	Q3199-09MS - 1	ELZ-25-000107MS	4.7061	17336	PPB	SMPL	-7470A	29 Sep 2025 12:23:08
U	Q3199-09MSD - 1	ELZ-25-000107MSD	4.5987	16940	PPB	SMPL	-7470A	29 Sep 2025 12:25:24
U	PB169872BL - 1	PBW	0.0883	307	PPB	SMPL	-7470A	29 Sep 2025 12:27:43
U	PB169872BS - 1	LCSW	3.5867	13208	PPB	SMPL	-7470A	29 Sep 2025 12:30:09
U	Q2893-09 - 1	MDL-WATER-03-QT3-2025	0.0500	166	PPB	SMPL	-7470A	29 Sep 2025 12:35:30
U	Q3172-01 - 1	FW7D	0.7206	2639	PPB	SMPL	-7470A	29 Sep 2025 12:37:46
U	Q3172-02 - 1	FW6D	0.0872	303	PPB	SMPL	-7470A	29 Sep 2025 12:40:03
U	Q3205-01 - 1	TRE-25-0106	0.0853	296	PPB	SMPL	-7470A	29 Sep 2025 12:42:21
U	Q3205-01DUP - 1	TRE-25-0106DUP	-0.0302	-130	PPB	SMPL	-7470A	29 Sep 2025 12:44:38
U	Q3205-01MS - 1	TRE-25-0106MS	3.8741	14268	PPB	SMPL	-7470A	29 Sep 2025 12:51:41
U	CCV55 - 1	CCV55	4.7533	17510	PPB	SMPL	-7470A	29 Sep 2025 12:53:56
U	CCB55 - 1	CCB55	-0.0587	-235	PPB	SMPL	-7470A	29 Sep 2025 12:56:16
U	Q3205-01MSD - 1	TRE-25-0106MSD	4.0236	14819	PPB	SMPL	-7470A	29 Sep 2025 12:58:34
U	PB169840TB - 1		-0.0397	-165	PPB	SMPL	-7470A	29 Sep 2025 13:00:50
U	PB169843TB - 1		0.0327	102	PPB	SMPL	-7470A	29 Sep 2025 13:03:08
U	Q3208-12LX5 - 1		0.0256	76	PPB	SMPL	-7470A	29 Sep 2025 13:08:16
U	Q3208-12A - 1		3.5875	13211	PPB	SMPL	-7470A	29 Sep 2025 13:10:31
U	Q3199-09LX5 - 1		-0.0376	-157	PPB	SMPL	-7470A	29 Sep 2025 13:12:48
U	Q3199-09A - 1		3.6956	13241	PPB	SMPL	-7470A	29 Sep 2025 13:15:06
U	Q3205-01LX5 - 1		-0.0048	-36	PPB	SMPL	-7470A	29 Sep 2025 13:17:24
U	Q3205-01A - 1		4.0111	14773	PPB	SMPL	-7470A	29 Sep 2025 13:19:44
U	CCV56 - 1	CCV56	4.8504	17868	PPB	SMPL	-7470A	29 Sep 2025 13:22:00
U	CCB56 - 1	CCB56	-0.0517	-209	PPB	SMPL	-7470A	29 Sep 2025 13:24:19
U	Q2893-07 - 1		0.0525	175	PPB	SMPL	-7470A	29 Sep 2025 13:26:38
U	Q2893-07RE - 1		0.1297	460	PPB	SMPL	-7470A	29 Sep 2025 13:34:20
U	CCV57 - 1	CCV57	4.8108	17722	PPB	SMPL	-7470A	29 Sep 2025 13:36:36
U	CCB57 - 1	CCB57	-0.0443	-182	PPB	SMPL	-7470A	29 Sep 2025 13:38:52