

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137348

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	09/26/2025	13:15
2	Q3170-02		1	11	100.00	>212.0	>212.0	09/26/2025	13:45
3	Q3170-02DUP		1	11	100.00	>212.0	>212.0	09/26/2025	14:15
4	Q3191-01		1	12	100.00	>212.0	>212.0	09/26/2025	14:45
5	Q3224-01		1	4	7.00	44.6	44.9	09/26/2025	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137348

WorkList Name : fp-09-26

WorkList ID : 192107

Department : Wet-Chemistry

Date : 09-26-2025 08:11:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3170-02	MOD-25-0271	Solid	Flash Point	Cool 4 deg C	PSEG03	J11	09/23/2025	1010B
Q3191-01	459	Solid	Flash Point	Cool 4 deg C	PSEG03	J41	09/24/2025	1010B
Q3224-01	60159	Solid	Flash Point	Cool 4 deg C	PSEG03	D31	09/26/2025	1010B

Date/Time 09/26/2025 13:00
 Raw Sample Received by: R14 WCC
 Raw Sample Relinquished by: JS WCC

Date/Time 09/26/2025 16:00
 Raw Sample Received by: JP WCC
 Raw Sample Relinquished by: R14 WCC