

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: Sohil

Run Number: LB137349

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	29.00	84.2	83.9	09/29/2025	09:45
2	Q3199-08		1	12	100.00	>212.0	>212.0	09/29/2025	10:15
3	Q3199-08DUP		1	13	100.00	>212.0	>212.0	09/29/2025	10:45
4	Q3199-09		1	12	100.00	>212.0	>212.0	09/29/2025	11:15
5	Q3205-01		1	12	100.00	>212.0	>212.0	09/29/2025	11:45
6	Q3224-02		1	13	100.00	>212.0	>212.0	09/29/2025	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB137349

WorkList Name : FLAS- POINT

WorkList ID : 192153

Department : Wet-Chemistry

Date : 09-29-2025 09:08:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3199-08	ELZ-25-000118	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/24/2025	1010B
Q3199-09	ELZ-25-000107	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/24/2025	1010B
Q3205-01	TRE-25-0106	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/25/2025	1010B
Q3224-02	60490	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/26/2025	1010B

Date/Time 09/29/25 09:30
 Raw Sample Received by: 12100
 Raw Sample Relinquished by: CP sm

Date/Time 09/29/25 12:55
 Raw Sample Received by: CP sm
 Raw Sample Relinquished by: 12100