

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB137354

Analyst By : JIGNESH

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/30/2025	06:20
2	CAL2	1	Water	NA	NA	20.2	7.00	09/30/2025	06:22
3	CAL3	1	Water	NA	NA	20.3	10.02	09/30/2025	06:25
4	ICV	1	Water	NA	NA	20.3	7.01	09/30/2025	06:30
5	CCV1	1	Water	NA	NA	20.3	2.01	09/30/2025	06:33
6	Q3224-02	1	Water	NA	NA	22.1	7.46	09/30/2025	06:40
7	Q3224-02DUP	1	Water	NA	NA	22.2	7.48	09/30/2025	06:42
8	Q3241-01	1	Water	NA	NA	20.3	7.03	09/30/2025	06:48
9	Q3241-01DUP	1	Water	NA	NA	20.4	7.05	09/30/2025	06:50
10	Q3241-02	1	Water	NA	NA	20.7	5.70	09/30/2025	07:00
11	Q3241-02DUP	1	Water	NA	NA	20.9	5.72	09/30/2025	07:02
12	Q3243-02	1	Water	NA	NA	20.1	6.35	09/30/2025	07:05
13	Q3243-02DUP	1	Water	NA	NA	20.2	6.37	09/30/2025	07:10
14	CCV2	1	Water	NA	NA	20.3	12.02	09/30/2025	07:15

WORKLIST(Hardcopy Internal Chain)

VB 134354

WorkList Name : ph q3246 WorkList ID : 192160 Department : Wet-Chemistry Date : 09-30-2025 06:06:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3224-02	60490	Water	pH	Cool 4 deg C	PSEG03	D31	09/26/2025	9040C
Q3241-01	92625	Water	pH	Cool 4 deg C	PSEG03	D21	09/29/2025	9040C
Q3241-02	92225	Water	pH	Cool 4 deg C	PSEG03	D21	09/29/2025	9040C
Q3243-02	91825C	Water	pH	Cool 4 deg C	PSEG03	D31	09/29/2025	9040C

Date/Time 09/30/25 06:15
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 09/30/25 12:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]