

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/3/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 10/02/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 104
Time OUT: 08:40
Out Date: 10/03/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % solids-oven

QC:LB137394

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q2409-05	COP-SOIL-PILE	12	1.18	10.58	11.76	11.38	96.4	
Q3249-03	S-3 (4-4.5)	1	1.11	10.45	11.56	9.51	80.4	
Q3249-08	S-8 (4.5-5)	2	1.19	10.75	11.94	10.38	85.5	
Q3249-16	S-14 (4-4.5)	3	1.13	10.32	11.45	9.34	79.6	
Q3249-19	S-18 (4.5-5)	4	1.18	10.64	11.82	10.8	90.4	
Q3249-22	S-20 (4-4.5)	5	1.18	10.47	11.65	10.64	90.4	
Q3249-24	S-22 (4.5-5)	6	1.19	10.11	11.3	10.33	90.4	
Q3249-27	S-24 (4.5-5)	7	1.15	11.05	12.2	9.82	78.5	
Q3269-01	EG1-TP1	8	1.18	10.78	11.96	10.97	90.8	
Q3269-02	EG1-TP1-E2	9	1.13	10.63	11.76	10.8	91.0	
Q3269-04	EG1-TP2	10	1.13	10.85	11.98	10.97	90.7	
Q3269-05	EG1-TP2-E2	11	1.11	10.77	11.88	11.13	93.0	
Q3270-01	SG-1-TEST-PIT	16	1.18	10.73	11.91	10.77	89.4	
Q3270-02	SG-1-TEST-PIT	17	1.19	10.32	11.51	10.79	93.0	
Q3271-01	72-11929	18	1.13	9.61	10.74	9.79	90.1	
Q3271-02	72-11929-E2	19	1.19	10.47	11.66	10.57	89.6	
Q3272-01	VNJ-235	20	1.19	10.68	11.87	10.15	83.9	
Q3272-02	VNJ-235-E2	21	1.13	10.47	11.6	10.15	86.2	
Q3272-03	R7	22	1.12	10.85	11.97	7.78	61.4	
Q3272-04	R7-E2	23	1.18	10.67	11.85	8.98	73.1	
Q3272-06	RT5417-CONCRETE	24	1.00	1.00	2.00	2.00	100.0	Concreate sample
Q3276-01	SB2	13	1.14	10.92	12.06	10.6	86.6	
Q3276-02	SB2A	14	1.15	10.38	11.53	10.6	91.0	
Q3277-01	WC1	15	1.18	10.41	11.59	11.34	97.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$