

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137398

Ambient Barometric Pressure (mmHg): 770.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.8	10/02/2025	10:20
2	Q3243-02		1	14	100.00	>212.0	>212.0	10/02/2025	11:10
3	Q3263-01		1	15	100.00	>212.0	>212.0	10/02/2025	11:40
4	Q3263-02		1	14	100.00	>212.0	>212.0	10/02/2025	12:45
5	Q3263-02DUP		1	14	100.00	>212.0	>212.0	10/02/2025	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

16137398

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-10-01

WorkList ID : 192238

Department : Wet-Chemistry

Date : 10-01-2025 16:10:26

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3243-02	91825C	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/29/2025	1010B
Q3263-01	251818	Water	Flash Point	Cool 4 deg C	PSEG03	D31	10/01/2025	1010B
Q3263-02	266380	Water	Flash Point	Cool 4 deg C	PSEG03	D31	10/01/2025	1010B

Date/Time 10/02/2025 10.10
Raw Sample Received by: RMCWS
Raw Sample Relinquished by: RMCWS

Date/Time 10/02/2025 17.20
Raw Sample Received by: RMCWS
Raw Sample Relinquished by: RMCWS