

## Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137400

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W3178         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W3093         |
| PH 10.01 BUFFER,COLOR CD 475ML | W3191         |
| buffer solution pH 7 yellow    | W3217         |
| Buffer Solution, PH2 (500ml)   | W3161         |
| pH 12.00 Buffer                | W3200         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 10/02/2025 | 15:00     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.00        | 10/02/2025 | 15:01     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.3             | 10.02       | 10/02/2025 | 15:05     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.3             | 7.01        | 10/02/2025 | 15:10     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 10/02/2025 | 15:11     |
| 6   | Q3256-04    | 1  | Solid  | 20.03       | 20          | 20.9             | 9.25        | 10/02/2025 | 15:15     |
| 7   | Q3256-08    | 1  | Solid  | 20.04       | 20          | 20.8             | 9.26        | 10/02/2025 | 15:22     |
| 8   | Q3262-03    | 1  | Solid  | 20.03       | 20          | 20.3             | 5.28        | 10/02/2025 | 15:30     |
| 9   | Q3262-06    | 1  | Solid  | 20.02       | 20          | 20.5             | 5.32        | 10/02/2025 | 15:35     |
| 10  | Q3262-09    | 1  | Solid  | 20.03       | 20          | 20.1             | 6.73        | 10/02/2025 | 15:40     |
| 11  | Q3269-03    | 1  | Solid  | 20.02       | 20          | 20.1             | 8.50        | 10/02/2025 | 15:44     |
| 12  | Q3269-06    | 1  | Solid  | 20.03       | 20          | 20.6             | 8.10        | 10/02/2025 | 15:47     |
| 13  | Q3269-06DUP | 1  | Solid  | 20.04       | 20          | 20.7             | 8.11        | 10/02/2025 | 15:50     |
| 14  | CCV2        | 1  | Water  | NA          | NA          | 20.3             | 12.02       | 10/02/2025 | 15:52     |