

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137404

Ambient Barometric Pressure (mmHg): 770.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.8	10/02/2025	10:20
2	Q3225-01		1	11	100.00	>212.0	>212.0	10/02/2025	15:45
3	Q3225-01DUP		1	11	100.00	>212.0	>212.0	10/02/2025	16:15
4	Q3244-01		1	9	92.00	197.6	197.0	10/02/2025	16:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB137404

WorkList Name : flash- point-

WorkList ID : 192237

Department : Wet-Chemistry

Date : 10-01-2025 09:10:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3225-01	60052	Solid	Flash Point	Cool 4 deg C	PSEG03	D31	09/26/2025	1010B
Q3244-01	80625	Solid	Flash Point	Cool 4 deg C	PSEG03	D31	09/29/2025	1010B

Date/Time 10/02/2025 10:10
 Raw Sample Received by: 1214 WSC
 Raw Sample Relinquished by: 1214 WSC

Date/Time 10/02/2025 17:20
 Raw Sample Received by: 1214 WSC
 Raw Sample Relinquished by: 1214 WSC