

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137411

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/03/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	7.00	10/03/2025	09:01
3	CAL3	1	Water	NA	NA	20.3	10.02	10/03/2025	09:05
4	ICV	1	Water	NA	NA	20.2	7.00	10/03/2025	09:08
5	CCV1	1	Water	NA	NA	20.2	2.02	10/03/2025	09:10
6	Q3263-01	1	Water	NA	NA	20.4	8.53	10/03/2025	09:20
7	Q3263-01DUP	1	Water	NA	NA	20.6	8.55	10/03/2025	09:22
8	Q3263-02	1	Water	NA	NA	20.4	7.60	10/03/2025	09:27
9	Q3263-02DUP	1	Water	NA	NA	20.6	7.62	10/03/2025	09:29
10	CCV2	1	Water	NA	NA	20.3	12.02	10/03/2025	09:30

WORKLIST(Hardcopy Internal Chain)

VB137411

WorkList Name : ph qq3263 WorkList ID : 192265 Department : Wet-Chemistry Date : 10-03-2025 08:39:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3263-01	251818	Water	pH	Cool 4 deg C	PSEG03	D31	10/01/2025	9040C
Q3263-02	266380	Water	pH	Cool 4 deg C	PSEG03	D31	10/01/2025	9040C

Date/Time 10/03/25 08:45
Raw Sample Received by: sp wpc,
Raw Sample Relinquished by: cd sm

Date/Time 10/03/25 12:35
Raw Sample Received by: cd sm
Raw Sample Relinquished by: sp wpc