

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137416

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3194

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	10/03/2025	12:10
2	Q3241-01		1	14	100.00	>212.0	>212.0	10/03/2025	12:40
3	Q3241-01DUP		1	14	100.00	>212.0	>212.0	10/03/2025	13:10
4	Q3241-02		1	15	100.00	>212.0	>212.0	10/03/2025	13:40
5	Q3252-04		1	15	100.00	>212.0	>212.0	10/03/2025	14:10
6	Q3252-05		1	14	100.00	>212.0	>212.0	10/03/2025	14:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137416

WorkList Name : FP-10-02

WorkList ID : 192259

Department : Wet-Chemistry

Date : 10-02-2025 18:07:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3241-01	92625	Water	Flash Point	Cool 4 deg C	PSEG03	D21	09/29/2025	1010B
Q3241-02	92225	Water	Flash Point	Cool 4 deg C	PSEG03	D21	09/29/2025	1010B
Q3252-04	AUD-25-0162-0164	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/30/2025	1010B
Q3252-05	AUD-25-0161	Water	Flash Point	Cool 4 deg C	PSEG03	D31	09/30/2025	1010B

Date/Time 10/03/2025 11:25
 Raw Sample Received by: R14 CWL
 Raw Sample Relinquished by: 78 (CWL)

Date/Time 10/03/2025 15:20
 Raw Sample Received by: 88 CWC
 Raw Sample Relinquished by: R14 CWC