

LB137429


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50031



A= 0.0000e+000

B= 1.9629e-004

C= 1.8288e-002

Rho= 0.9995032

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.037	-0.037	-280	0.000	-280				
0.2	0.200	0.188	-0.012	867	0.0 %	867				
2.5	2.500	2.531	0.031	12801	0.0 %	12801				
5.0	5.000	4.953	-0.047	25140	0.0 %	25140				
7.5	7.500	7.725	0.225	39261	0.0 %	39261				
10.0	10.000	9.839	-0.161	50031	0.0 %	50031				

LB137429 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Oct 2025 12:19:59

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	-280	PPB	Std	0.0000	7471B	06 Oct 2025 14:39:49
S	0.2 - 1	50.2	-	867	PPB	Std	0.2000	7471B	06 Oct 2025 14:42:47
S	2.5 - 1	52.5	-	12801	PPB	Std	2.5000	7471B	06 Oct 2025 14:45:06
S	5.0 - 1	55	-	25140	PPB	Std	5.0000	7471B	06 Oct 2025 14:47:23
S	7.5 - 1	57.5	-	39261	PPB	Std	7.5000	7471B	06 Oct 2025 14:49:59
S	10.0 - 1	57.5	-	50031	PPB	Std	10.0000	7471B	06 Oct 2025 14:52:42
U	ICV59 - 1	ICV59	4.0696	20639	PPB	SMPL	-7471B		06 Oct 2025 14:56:06
U	ICB59 - 1	ICB59	-0.0494	-345	PPB	SMPL	-7471B		06 Oct 2025 14:58:33
U	CCV83 - 1	CCV83	4.8948	24843	PPB	SMPL	-7471B		06 Oct 2025 15:01:13
U	CCB83 - 1	CCB83	-0.0459	-327	PPB	SMPL	-7471B		06 Oct 2025 15:03:28
U	CRA - 1	CRA	0.1961	906	PPB	SMPL	-7471B		06 Oct 2025 15:06:09
U	HighStd - 1	HighStd	9.8012	49838	PPB	SMPL	-7471B		06 Oct 2025 15:08:24
U	ChkStd - 1	ChkStd	6.7849	34472	PPB	SMPL	-7471B		06 Oct 2025 15:10:39
U	PB169993BL - 1	PBS	-0.0473	-334	PPB	SMPL	-7471B		06 Oct 2025 15:13:31
U	PB169993BS - 1	LCSS	3.6421	18461	PPB	SMPL	-7471B		06 Oct 2025 15:16:19
U	Q3270-01 - 1	SG-1-TEST-PIT	0.3196	1535	PPB	SMPL	-7471B		06 Oct 2025 15:18:35
U	Q3271-01 - 1	72-11929	0.3561	1721	PPB	SMPL	-7471B		06 Oct 2025 15:21:15
U	Q3272-01 - 1	VNJ-235	6.9398	35261	PPB	SMPL	-7471B		06 Oct 2025 15:23:34
U	Q3272-03 - 1	R7	0.8629	4303	PPB	SMPL	-7471B		06 Oct 2025 15:25:52
U	Q3277-01 - 1	WC1	4.3063	21845	PPB	SMPL	-7471B		06 Oct 2025 15:28:36
U	CCV84 - 1	CCV84	4.8351	24539	PPB	SMPL	-7471B		06 Oct 2025 15:30:53
U	CCB84 - 1	CCB84	-0.0339	-266	PPB	SMPL	-7471B		06 Oct 2025 15:33:34
U	Q3288-01 - 1	TR-01-100325	0.3816	1851	PPB	SMPL	-7471B		06 Oct 2025 15:36:16
U	Q3289-01 - 1	VNJ-240	0.3567	1724	PPB	SMPL	-7471B		06 Oct 2025 15:38:31
U	Q3289-03 - 1	VNJ-261	0.5457	2687	PPB	SMPL	-7471B		06 Oct 2025 15:40:47
U	Q3289-05 - 1	72-11966	0.1123	479	PPB	SMPL	-7471B		06 Oct 2025 15:43:03
U	Q3290-01 - 1	CL-02-100325	0.2868	1368	PPB	SMPL	-7471B		06 Oct 2025 15:45:18
U	Q3290-01DUP - 1	CL-02-100325DUP	0.3180	1527	PPB	SMPL	-7471B		06 Oct 2025 15:47:35
U	Q3290-01MS - 1	CL-02-100325MS	3.9722	20143	PPB	SMPL	-7471B		06 Oct 2025 15:49:51
U	Q3290-01MSD - 1	CL-02-100325MSD	4.2590	21604	PPB	SMPL	-7471B		06 Oct 2025 15:52:08
U	Q3290-01LX5 - 1		0.0273	46	PPB	SMPL	-7471B		06 Oct 2025 15:54:48
U	Q3290-01A - 1		4.0403	20490	PPB	SMPL	-7471B		06 Oct 2025 15:57:27
U	CCV85 - 1	CCV85	4.8338	24532	PPB	SMPL	-7471B		06 Oct 2025 15:59:43
U	CCB85 - 1	CCB85	-0.0343	-268	PPB	SMPL	-7471B		06 Oct 2025 16:02:23