

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB137432

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	29.00	84.2	83.9	10/06/2025	09:25
2	Q3287-02		1	9	61.00	141.8	141.5	10/06/2025	09:55
3	Q3287-02DUP		1	9	60.00	140.0	139.7	10/06/2025	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LD137432

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-10-06-

WorkList ID : 192298

Department : Wet-Chemistry

Date : 10-06-2025 08:15:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3287-02	TSS-25-0001-0002	Solid	Flash Point	Cool 4 deg C	PSEG03	D31	10/03/2025	1010B

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/06/25 09:10
12/507
2060017

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

10/06/25 11:10
2060017
12/507