

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137450

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/08/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	7.00	10/08/2025	09:01
3	CAL3	1	Water	NA	NA	20.2	10.02	10/08/2025	09:03
4	ICV	1	Water	NA	NA	20.3	7.00	10/08/2025	09:05
5	CCV1	1	Water	NA	NA	20.3	2.01	10/08/2025	09:08
6	Q3305-01	1	Solid	20.02	20	22.8	6.87	10/08/2025	09:25
7	Q3305-01DUP	1	Solid	20.03	20	22.9	6.89	10/08/2025	09:28
8	CCV2	1	Water	NA	NA	20.2	12.02	10/08/2025	09:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s q3305 WorkList ID : 192337 Department : Wet-Chemistry Date : 10-08-2025 08:32:56

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3305-01	AU-06-100725	Solid	pH	Cool 4 deg C	PSEG05	D31	10/07/2025	9045D

Date/Time 10/08/25 08:45
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11:15
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]