

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137451

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/08/2025	09:00
2	CAL2	1	Water	NA	NA	20.2	7.00	10/08/2025	09:01
3	CAL3	1	Water	NA	NA	20.2	10.02	10/08/2025	09:03
4	ICV	1	Water	NA	NA	20.3	7.00	10/08/2025	09:05
5	CCV1	1	Water	NA	NA	20.2	2.02	10/08/2025	09:37
6	Q3293-02	1	Solid	20.04	20	22.3	13.5	10/08/2025	09:50
7	Q3302-06	1	Solid	20.02	20	23.5	5.55	10/08/2025	10:00
8	Q3302-06DUP	1	Solid	20.03	20	23.6	5.56	10/08/2025	10:01
9	CCV2	1	Water	NA	NA	20.2	12.02	10/08/2025	10:05

WORKLIST(Hardcopy Internal Chain)

VP 137451

WorkList Name : corrosivity q3302 WorkList ID : 192336 Department : Wet-Chemistry Date : 10-08-2025 08:32:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3293-02	CONC-STORAGE-PAD-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/06/2025	9045D
Q3302-06	OR-658-COMP-01	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/07/2025	9045D

Date/Time 10/08/25 08:45
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10/08/25 11:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]