

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137470

Slope : 98.5

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/09/2025	08:20
2	CAL2	1	Water	NA	NA	20.3	7.00	10/09/2025	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	10/09/2025	08:17
4	ICV	1	Water	NA	NA	20.3	7.02	10/09/2025	08:20
5	CCV1	1	Water	NA	NA	20.1	2.01	10/09/2025	08:40
6	Q3315-02	1	Solid	20.02	20	21.6	5.62	10/09/2025	08:50
7	Q3315-02DUP	1	Solid	20.03	20	21.7	5.64	10/09/2025	08:52
8	Q3318-03	1	Solid	20.02	20	21.6	7.33	10/09/2025	09:00
9	CCV2	1	Water	NA	NA	20.3	12.02	10/09/2025	09:05