

LB137486


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37863



A= 0.0000e+000

B= 2.6628e-004

C= 3.5758e-003

Rho= 0.9998331

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.080	0.080	286	0.000	286	0			
0.2	0.200	0.204	0.004	753	0.0 %	753	0			
2.5	2.500	2.476	-0.024	9286	0.0 %	9286	0			
5.0	5.000	4.893	-0.107	18362	0.0 %	18362				
7.5	7.500	7.461	-0.039	28005	0.0 %	28005				
10.0	10.000	10.086	0.086	37863	0.0 %	37863				

LB137486 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Oct 2025 10:31:11

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	286	PPB	Std	0.0000	7470A	10 Oct 2025 10:49:29
S	0.2 - 1	50.2	-	753	PPB	Std	0.2000	7470A	10 Oct 2025 10:51:47
S	2.5 - 1	52.5	-	9286	PPB	Std	2.5000	7470A	10 Oct 2025 10:54:06
S	5.0 - 1	55	-	18362	PPB	Std	5.0000	7470A	10 Oct 2025 10:56:23
S	7.5 - 1	57.5	-	28005	PPB	Std	7.5000	7470A	10 Oct 2025 10:58:41
S	10.0 - 1	510	-	37863	PPB	Std	10.0000	7470A	10 Oct 2025 11:01:00
U	ICV67 - 1	ICV67	3.8738	14534	PPB	SMPL	-	7470A	10 Oct 2025 11:04:04
U	ICB67 - 1	ICB67	-0.0148	-69	PPB	SMPL	-	7470A	10 Oct 2025 11:06:20
U	CCV16 - 1	CCV16	4.9975	18754	PPB	SMPL	-	7470A	10 Oct 2025 11:11:33
U	CCB16 - 1	CCB16	-0.0249	-107	PPB	SMPL	-	7470A	10 Oct 2025 11:13:47
U	CRA - 1	CRA	0.1865	687	PPB	SMPL	-	7470A	10 Oct 2025 11:16:06
U	HighStd - 1	HighStd	10.0782	37834	PPB	SMPL	-	7470A	10 Oct 2025 11:18:21
U	ChkStd - 1	ChkStd	6.6918	25117	PPB	SMPL	-	7470A	10 Oct 2025 11:20:37
U	PB170048BL - 1	PBW	0.0379	129	PPB	SMPL	-	7470A	10 Oct 2025 11:22:59
U	PB170048BS - 1	LCSW	3.8884	14589	PPB	SMPL	-	7470A	10 Oct 2025 11:25:20
U	Q3315-02 - 1	314	0.1162	423	PPB	SMPL	-	7470A	10 Oct 2025 11:27:36
U	Q3315-02DUP - 1	314DUP	0.1338	489	PPB	SMPL	-	7470A	10 Oct 2025 11:29:56
U	Q3315-02MS - 1	314MS	4.0484	15190	PPB	SMPL	-	7470A	10 Oct 2025 11:32:13
U	Q3315-02MSD - 1	314MSD	4.0841	15324	PPB	SMPL	-	7470A	10 Oct 2025 11:34:30
U	PB170056BL - 1	PBW	0.0600	212	PPB	SMPL	-	7470A	10 Oct 2025 11:36:48
U	CCV17 - 1	CCV17	4.8430	18174	PPB	SMPL	-	7470A	10 Oct 2025 11:39:06
U	CCB17 - 1	CCB17	-0.0337	-140	PPB	SMPL	-	7470A	10 Oct 2025 11:41:21
U	PB170056BS - 1	LCSW	3.8282	14363	PPB	SMPL	-	7470A	10 Oct 2025 11:43:39
U	Q3318-04 - 1	3436	0.2712	1005	PPB	SMPL	-	7470A	10 Oct 2025 11:45:54
U	Q3318-04DUP - 1	3436DUP	0.2733	1013	PPB	SMPL	-	7470A	10 Oct 2025 11:48:13
U	Q3318-04MS - 1	3436MS	2.6789	10047	PPB	SMPL	-	7470A	10 Oct 2025 11:50:28
U	Q3318-04MSD - 1	3436MSD	2.4494	9185	PPB	SMPL	-	7470A	10 Oct 2025 11:52:44
U	PB170027B - 1	PB170027B	-0.0111	-55	PPB	SMPL	-	7470A	10 Oct 2025 11:55:03
U	Q3315-02LX5 - 1		0.0041	2	PPB	SMPL	-	7470A	10 Oct 2025 11:57:23
U	Q3315-02A - 1		3.8346	14387	PPB	SMPL	-	7470A	10 Oct 2025 11:59:40
U	Q3318-04LX5 - 1		0.0451	156	PPB	SMPL	-	7470A	10 Oct 2025 12:01:57
U	Q3318-04A - 1		3.6061	13529	PPB	SMPL	-	7470A	10 Oct 2025 12:04:15
U	CCV18 - 1	CCV18	4.6588	17482	PPB	SMPL	-	7470A	10 Oct 2025 12:06:30
U	CCB18 - 1	CCB18	-0.0108	-54	PPB	SMPL	-	7470A	10 Oct 2025 12:08:48