

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137493

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	10/10/2025	11:35
2	Q3318-03		1	4	7.00	44.6	43.9	10/10/2025	12:05
3	Q3318-03DUP		1	4	7.00	44.6	43.9	10/10/2025	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6137493

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-3318

WorkList ID : 192397

Department : Wet-Chemistry

Date : 10-10-2025 08:13:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3318-03	3435	Solid	Flash Point	Cool 4 deg C	PSEG03	D31	10/08/2025	1010B

Date/Time 10/10/2025 11:25
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: RM CWG

Date/Time 10/10/2025 14:15
Raw Sample Received by: RM CWG
Raw Sample Relinquished by: RM CWG