

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB137503

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/13/2025	09:45
2	CAL2	1	Water	NA	NA	20.3	7.00	10/13/2025	09:46
3	CAL3	1	Water	NA	NA	20.3	10.02	10/13/2025	09:47
4	ICV	1	Water	NA	NA	20.2	7.01	10/13/2025	09:48
5	CCV1	1	Water	NA	NA	20.2	2.01	10/13/2025	09:50
6	Q3330-01	1	Water	NA	NA	20.1	7.92	10/13/2025	10:00
7	Q3330-01DUP	1	Water	NA	NA	20.3	7.93	10/13/2025	10:03
8	Q3331-04	1	Water	NA	NA	20.8	6.32	10/13/2025	10:05
9	Q3331-04DUP	1	Water	NA	NA	20.8	6.35	10/13/2025	10:10
10	Q3332-02	1	Water	NA	NA	20.6	7.94	10/13/2025	10:15
11	Q3332-02DUP	1	Water	NA	NA	20.8	7.96	10/13/2025	10:19
12	Q3332-03	1	Water	NA	NA	20.2	10.11	10/13/2025	10:30
13	Q3332-03DUP	1	Water	NA	NA	20.3	10.12	10/13/2025	10:33
14	Q3334-02	1	Water	NA	NA	20.1	7.79	10/13/2025	10:37
15	Q3334-02DUP	1	Water	NA	NA	20.3	7.82	10/13/2025	10:39
16	CCV2	1	Water	NA	NA	20.3	12.02	10/13/2025	10:40