

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137508

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	10/13/2025	12:00
2	Q3332-02		1	15	100.00	>212.0	>212.0	10/13/2025	12:30
3	Q3332-02DUP		1	15	100.00	>212.0	>212.0	10/13/2025	13:00
4	Q3332-03		1	14	100.00	>212.0	>212.0	10/13/2025	13:30
5	Q3334-02		1	15	100.00	>212.0	>212.0	10/13/2025	14:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06