


LB137509

7471B

INSTRUMENT ID: CV1

 μ Abs.:

51123



Linear

A= 0.0000e+000

B= 1.9695e-004

C= -5.0279e-003

Rho= 0.9996627

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	58	0.000	58				
0.2	0.200	0.211	0.011	1096	0.0 %	1096				
2.5	2.500	2.615	0.115	13301	0.0 %	13301				
5.0	5.000	4.806	-0.194	24426	0.0 %	24426				
7.5	7.500	7.499	-0.001	38100	0.0 %	38100				
10.0	10.000	10.064	0.064	51123	0.0 %	51123				

LB137509 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Oct 2025 12:55:21

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	58	PPB	Std	0.0000	7471B	13 Oct 2025 13:27:59
S	0.2 - 1	50.2	-	1096	PPB	Std	0.2000	7471B	13 Oct 2025 13:30:21
S	2.5 - 1	52.5	-	13301	PPB	Std	2.5000	7471B	13 Oct 2025 13:32:38
S	5.0 - 1	55	-	24426	PPB	Std	5.0000	7471B	13 Oct 2025 13:34:55
S	7.5 - 1	57.5	-	38100	PPB	Std	7.5000	7471B	13 Oct 2025 13:37:29
S	10.0 - 1	510	-	51123	PPB	Std	10.0000	7471B	13 Oct 2025 13:43:23
U	ICV68 - 1	ICV68	4.1122	20905	PPB	SMPL	-7471B		13 Oct 2025 13:49:00
U	ICB68 - 1	ICB68	-0.0484	-220	PPB	SMPL	-7471B		13 Oct 2025 13:51:25
U	CCV19 - 1	CCV19	4.8029	24412	PPB	SMPL	-7471B		13 Oct 2025 13:54:03
U	CCB19 - 1	CCB19	-0.0484	-220	PPB	SMPL	-7471B		13 Oct 2025 13:56:18
U	CRA - 1	CRA	0.2035	1059	PPB	SMPL	-7471B		13 Oct 2025 13:58:59
U	HighStd - 1	HighStd	10.1200	51409	PPB	SMPL	-7471B		13 Oct 2025 14:01:14
U	ChkStd - 1	ChkStd	6.9920	35527	PPB	SMPL	-7471B		13 Oct 2025 14:03:29
U	PB170072BL - 1	PBS	-0.0639	-299	PPB	SMPL	-7471B		13 Oct 2025 14:06:19
U	PB170072BS - 1	LCSS	3.5088	17841	PPB	SMPL	-7471B		13 Oct 2025 14:11:42
U	Q3321-01 - 1	P001-DS-01	4.4323	22530	PPB	SMPL	-7471B		13 Oct 2025 14:14:05
U	Q3322-01 - 1	OR-02-100925	0.4462	2291	PPB	SMPL	-7471B		13 Oct 2025 14:16:44
U	Q3323-01 - 1	SED-01-0-2	5.5645	28279	PPB	SMPL	-7471B		13 Oct 2025 14:19:27
U	Q3323-02 - 1	SED-02-0-2	1.7356	8838	PPB	SMPL	-7471B		13 Oct 2025 14:21:45
U	Q3323-03 - 1	SED-03-0-2	4.8244	24521	PPB	SMPL	-7471B		13 Oct 2025 14:24:34
U	CCV20 - 1	CCV20	4.7068	23924	PPB	SMPL	-7471B		13 Oct 2025 14:27:06
U	CCB20 - 1	CCB20	-0.0610	-284	PPB	SMPL	-7471B		13 Oct 2025 14:29:49
U	Q3323-04 - 1	SED-04-0-2	4.8998	24904	PPB	SMPL	-7471B		13 Oct 2025 14:32:29
U	Q3323-05 - 1	SED-05-0-2	3.1182	15858	PPB	SMPL	-7471B		13 Oct 2025 14:34:45
U	Q3323-06 - 1	SED-06-0-2	1.7736	9031	PPB	SMPL	-7471B		13 Oct 2025 14:37:27
U	Q3323-07 - 1	SED-07-0-2	8.4136	42745	PPB	SMPL	-7471B		13 Oct 2025 14:40:04
U	Q3326-01 - 1	TP-1	-0.0168	-60	PPB	SMPL	-7471B		13 Oct 2025 14:42:35
U	Q3326-05 - 1	TP-2	0.3300	1701	PPB	SMPL	-7471B		13 Oct 2025 14:45:24
U	Q3327-01 - 1	HADLEY-ROAD-TP	0.1764	921	PPB	SMPL	-7471B		13 Oct 2025 14:47:41
U	Q3327-01DUP - 1	HADLEY-ROAD-TPDUP	0.1673	875	PPB	SMPL	-7471B		13 Oct 2025 14:49:58
U	Q3327-01MS - 1	HADLEY-ROAD-TPMS	3.5289	17943	PPB	SMPL	-7471B		13 Oct 2025 14:52:15
U	Q3327-01MSD - 1	HADLEY-ROAD-TPMSD	3.6982	18803	PPB	SMPL	-7471B		13 Oct 2025 14:54:30
U	CCV21 - 1	CCV21	4.7403	24094	PPB	SMPL	-7471B		13 Oct 2025 14:57:08
U	CCB21 - 1	CCB21	-0.0505	-231	PPB	SMPL	-7471B		13 Oct 2025 14:59:45
U	Q3332-01 - 1	100225-D	0.5683	2911	PPB	SMPL	-7471B		13 Oct 2025 15:02:26
U	Q3327-01LX5 - 1		-0.0017	17	PPB	SMPL	-7471B		13 Oct 2025 15:04:41
U	Q3327-01A - 1		3.5681	18142	PPB	SMPL	-7471B		13 Oct 2025 15:06:58
U	CCV22 - 1	CCV22	4.7141	23961	PPB	SMPL	-7471B		13 Oct 2025 15:09:14
U	CCB22 - 1	CCB22	-0.0576	-267	PPB	SMPL	-7471B		13 Oct 2025 15:11:52