

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137510

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	10/14/2025	08:30
2	Q3330-01		1	5	28.00	82.4	82.4	10/14/2025	09:00
3	Q3331-04		1	15	100.00	>212.0	>212.0	10/14/2025	09:30
4	Q3331-04DUP		1	15	100.00	>212.0	>212.0	10/14/2025	10:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137510

WorkList Name : fp-10-13-

WorkList ID : 192427

Department : Wet-Chemistry

Date : 10-13-2025 12:55:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3330-01	2047	Water	Flash Point	Cool 4 deg C	PSEG03	D41	10/10/2025	1010B
Q3331-04	RINGWOOD	Water	Flash Point	Cool 4 deg C	PSEG03	D41	10/10/2025	1010B

Date/Time 10/14/2025 08:15
 Raw Sample Received by: RM (ws)
 Raw Sample Relinquished by: JPCOO

Date/Time 10/14/2025 10:45
 Raw Sample Received by: JPCOO
 Raw Sample Relinquished by: RM (ws)