

LB 137514


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39739



A= 0.0000e+000

B= 2.5162e-004

C= -3.7999e-002

Rho= 0.9995115

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.053	-0.053	-60	0.000	-60	0			
0.2	0.200	0.176	-0.024	851	0.0 %	851	0			
2.5	2.500	2.488	-0.012	10038	0.0 %	10038	0			
5.0	5.000	5.246	0.246	21002	0.0 %	21002				
7.5	7.500	7.382	-0.118	29488	0.0 %	29488				
10.0	10.000	9.961	-0.039	39739	0.0 %	39739				

lb137514 instrument id : cv1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Oct 2025 10:32:29

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Strnd Conc	Method	Date
S	0.0 - 1	50	-	-60	PPB	Std	0.0000	7470A	14 Oct 2025 10:42:16
S	0.2 - 1	50.2	-	851	PPB	Std	0.2000	7470A	14 Oct 2025 10:49:32
S	2.5 - 1	52.5	-	10038	PPB	Std	2.5000	7470A	14 Oct 2025 10:51:48
S	5.0 - 1	55	-	21002	PPB	Std	5.0000	7470A	14 Oct 2025 10:54:05
S	7.5 - 1	57.5	-	29488	PPB	Std	7.5000	7470A	14 Oct 2025 10:56:24
S	10.0 - 1	59	-	39739	PPB	Std	10.0000	7470A	14 Oct 2025 10:58:45
U	ICV70 - 1	ICV70	4.3412	17404	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:07:06
U	ICB70 - 1	ICB70	-0.0315	26	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:12:09
U	CCV26 - 1	CCV26	4.9216	19711	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:17:26
U	CCB26 - 1	CCB26	-0.0272	43	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:22:38
U	CRA - 1	CRA	0.1744	844	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:24:52
U	HIGH Std - 1	HIGH Std	9.6498	38502	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:27:07
U	CHK Std - 1	CHK Std	6.9046	27592	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:32:16
U	PB170073BL - 1	PBW	-0.0342	15	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:37:21
U	PB170073BS - 1	LCSW	4.1995	16841	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:39:37
U	Q332102 - 1	P001-05-01	-0.0722	-136	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:41:53
U	Q3326-04 - 1	TP-1	-0.0317	25	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:44:15
U	Q3326-08 - 1	TP-2	-0.0307	29	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:46:32
U	Q3327-02 - 1	HADLEY ROAD	-0.0239	56	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:48:49
U	Q3327-02DUP - 1	HADLEY ROADDUP	-0.0337	17	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:51:05
U	CCV27 - 1	CCV27	4.7098	18869	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:53:20
U	CCB27 - 1	CCB27	-0.0984	-240	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:55:35
U	Q3327-02MS - 1	HADLEY ROADMS	4.7382	18982	PPB	SMPL	-7470A	-7470A	14 Oct 2025 11:57:54
U	Q3327-02MSD - 1	HADLEY ROADMSD	3.8643	15509	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:00:10
U	PB170084BL - 1	PBW	-0.0893	-204	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:02:31
U	PB170084BS - 1	LCSW	4.1602	16685	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:04:50
U	Q3331-04 - 1	RINGWOOD	-0.0093	114	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:07:06
U	Q3331-04DUP - 1	RINGWOODDUP	-0.0138	96	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:09:26
U	Q3331-04MS - 1	RINGWOODMS	3.7803	15175	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:11:43
U	Q3331-04MSD - 1	RINGWOODMSD	3.8080	15285	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:18:24
U	Q3332-02 - 1	100725-B	-0.0805	-169	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:20:41
U	CCV28 - 1	CCV28	5.1081	20452	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:22:59
U	CCB28 - 1	CCB28	-0.1012	-251	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:25:15
U	PB170062TB - 1	PB170062TB	-0.0327	21	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:27:35
U	Q3327-02LX5 - 1		-0.0279	40	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:29:51
U	Q3327-02A - 1		4.2272	16951	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:32:06
U	Q3331-04LX5 - 1		-0.0823	-176	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:34:22
U	Q3331-04A - 1		4.4544	17854	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:36:41
U	CCV29 - 1	CCV29	5.0535	20235	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:38:57
U	CCB29 - 1	CCB29	-0.1125	-296	PPB	SMPL	-7470A	-7470A	14 Oct 2025 12:41:17