


Instrument: IC-1										Analyst: RM		Method: 300.0 / 9056A		Initial wt/Final Analyst	
Ident	Con F-	Con CL-	Con NO2	Con BR-	Con NO3	Con HPO4	Con SO4	Method name	date time						
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	IC1-100925	10/9/2025 10:46					10 RM/IZ	
STD2	0.4400	0.6460	0.6520	2.1410	0.5410	1.1180	3.2990	IC1-100925	10/9/2025 11:08					10 RM/IZ	
STD3	0.8130	1.2250	1.2220	4.0730	1.0190	1.9470	6.0910	IC1-100925	10/9/2025 11:29					10 RM/IZ	
STD4	0.9930	1.4970	1.4960	4.9940	1.2490	2.6690	7.5070	IC1-100925	10/9/2025 11:51					10 RM/IZ	
STD5	1.9540	2.9230	2.9220	9.7730	2.4350	4.8280	14.5570	IC1-100925	10/9/2025 12:12					10 RM/IZ	
STD6	3.9280	5.9230	5.9200	19.7520	4.9340	9.5170	29.5100	IC1-100925	10/9/2025 12:34					10 RM/IZ	
STD7	5.0720	7.5850	7.5890	25.2680	6.3230	12.9200	37.5370	IC1-100925	10/9/2025 12:55					10 RM/IZ	
ICV	1.9700	2.9180	2.9420	9.8420	2.3980	5.0030	14.7020	IC1-100925	10/9/2025 13:16					10 RM/IZ	
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	IC1-100925	10/9/2025 13:38					10 RM/IZ	
CCV	2.0370	3.0170	3.0250	10.1630	2.4570	4.8590	15.0070	IC1-100925	10/15/2025 9:11					10 RM/IZ	
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	IC1-100925	10/15/2025 9:32					10 RM/IZ	
LB137519BLS	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	IC1-100925	10/15/2025 9:54	5.00g/100mL	RM/IZ				
LB137519BSS	2.0390	3.0480	3.0280	10.1980	2.4820	5.1990	15.1320	IC1-100925	10/15/2025 10:15	5.00g/100mL	RM/IZ				
Q3341-01	0.3140	29.6780	0.0000	0.0000	0.0850	0.2620	1.0890	IC1-100925	10/15/2025 10:37	5.04g/100mL	RM/IZ				
Q3341-01MS	2.0810	31.7580	3.0240	10.1310	2.4910	4.2830	15.5330	IC1-100925	10/15/2025 10:58	5.01g/100mL	RM/IZ				
Q3341-01MSD	2.0710	31.7220	3.0220	10.1270	2.4930	4.2820	15.5290	IC1-100925	10/15/2025 11:20	5.03g/100mL	RM/IZ				
Q3341-02	0.1950	1.8200	0.0000	0.0000	0.0000	0.0000	0.8400	IC1-100925	10/15/2025 11:42	5.03g/100mL	RM/IZ				
Q3341-03	0.0900	1.5500	0.0000	0.0000	0.0000	0.4310	1.2110	IC1-100925	10/15/2025 12:03	5.00g/100mL	RM/IZ				
Q3341-01DLX10	0.0670	2.6860	0.0000	0.0000	0.0000	0.0000	0.4940	IC1-100925	10/15/2025 12:25	5.04g/100mL	RM/IZ				
CCV	2.0310	3.0200	3.0260	10.1750	2.4740	4.9560	15.0450	IC1-100925	10/15/2025 12:46					10 RM/IZ	
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	IC1-100925	10/15/2025 13:08					10 RM/IZ	

Clear table

Instrument ID: IC-2 Analyst: IZ Method: 300.0 / 9056A											
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000 IC1-100925	10/9/2025 10:46	Final	10 RM/IZ
STD2	0.4400	0.6460	0.6520	2.1410	0.5410	1.1180	3.2990	IC1-100925	10/9/2025 11:08		10 RM/IZ
STD3	0.8130	1.2250	1.2220	4.0730	1.0190	1.9470	6.0910	IC1-100925	10/9/2025 11:29		10 RM/IZ
STD4	0.9930	1.4970	1.4960	4.9940	1.2490	2.6690	7.5070	IC1-100925	10/9/2025 11:51		10 RM/IZ
STD5	1.9540	2.9230	2.9220	9.7730	2.4350	4.8280	14.5570	IC1-100925	10/9/2025 12:12		10 RM/IZ
STD6	3.9280	5.9230	5.9200	19.7520	4.9340	9.5170	29.5100	IC1-100925	10/9/2025 12:34		10 RM/IZ
STD7	5.0720	7.5850	7.5890	25.2680	6.3230	12.9200	37.5370	IC1-100925	10/9/2025 12:55		10 RM/IZ
ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4				
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000				
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000				
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000				
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000				
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000				
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000				
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000				
ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4				
STD1											
STD2	10.0000	7.6667	8.6667	7.0500	8.2000	11.8000	9.9667				
STD3	1.6250	2.0833	1.8333	1.8250	1.9000	-2.6500	1.5167				
STD4	-0.7000	-0.2000	-0.2667	-0.1200	-0.0800	6.7600	0.0933				
STD5	-2.3000	-2.5667	-2.6000	-2.2700	-2.6000	-3.4400	-2.9533				
STD6	-1.8000	-1.2833	-1.3333	-1.2400	-1.3200	-4.8300	-1.6333				
STD7	1.4400	1.1333	1.1867	1.0720	1.1680	3.3600	1.4514				

Fluoride (Anions)




Function: $A = -3.86859E-3 + 0.0146255 \times Q$

Relative standard deviation 2.744994 %

Correlation coefficient 0.999599

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	0.400	10.0	1.0	1.0	0.061	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	0.800	10.0	1.0	1.0	0.115	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	1.000	10.0	1.0	1.0	0.141	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	2.000	10.0	1.0	1.0	0.282	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	4.000	10.0	1.0	1.0	0.571	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	5.000	10.0	1.0	1.0	0.738	STD7	2025-10-09 12:55:27 UTC-4	used

Chloride (Anions)



Function: $A = -5.87555E-3 + 0.0101528 \times Q$


Relative standard deviation 2.276543 %

Correlation coefficient 0.999727

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	0.600	10.0	1.0	1.0	0.060	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	1.200	10.0	1.0	1.0	0.119	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	1.500	10.0	1.0	1.0	0.146	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	3.000	10.0	1.0	1.0	0.291	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	6.000	10.0	1.0	1.0	0.596	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	7.500	10.0	1.0	1.0	0.764	STD7	2025-10-09 12:55:27 UTC-4	used

Nitrite (Anions)

($\mu\text{S/cm}$) $\times$ min



Function: $A = -0.0165604 + 0.0226258 \times Q$


Relative standard deviation 2.379226 %

Correlation coefficient 0.999705

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	0.600	10.0	1.0	1.0	0.131	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	1.200	10.0	1.0	1.0	0.260	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	1.500	10.0	1.0	1.0	0.322	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	3.000	10.0	1.0	1.0	0.645	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	6.000	10.0	1.0	1.0	1.323	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	7.500	10.0	1.0	1.0	1.700	STD7	2025-10-09 12:55:27 UTC-4	used

Bromide (Anions)

($\mu\text{S}/\text{cm}$) $\times$ min



Function: $A = -6.89623E-3 + 4.37562E-3 \times Q$


Relative standard deviation: 2.114692 %

Correlation coefficient: 0.999763

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	2.000	10.0	1.0	1.0	0.087	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	4.000	10.0	1.0	1.0	0.171	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	5.000	10.0	1.0	1.0	0.212	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	10.000	10.0	1.0	1.0	0.421	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	20.000	10.0	1.0	1.0	0.857	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	25.000	10.0	1.0	1.0	1.099	STD7	2025-10-09 12:55:27 UTC-4	used

Nitrate (Anions)

($\mu\text{S}/\text{cm}$) $\times$ min



Function: $A = -0.0141807 + 0.0252306 \times Q$


Relative standard deviation: 2.346972 %

Correlation coefficient 0.999712

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	0.500	10.0	1.0	1.0	0.122	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	1.000	10.0	1.0	1.0	0.243	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	1.250	10.0	1.0	1.0	0.301	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	2.500	10.0	1.0	1.0	0.600	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	5.000	10.0	1.0	1.0	1.231	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	6.250	10.0	1.0	1.0	1.581	STD7	2025-10-09 12:55:27 UTC-4	used

Phosphate (Anions)

($\mu\text{S/cm}$) $\times$ min




Function: $A = -0.0117521 + 8.50315E-3 \times Q$

Relative standard deviation 6.489714 %

Correlation coefficient 0.997825

Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	1.000	10.0	1.0	1.0	0.083	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	2.000	10.0	1.0	1.0	0.154	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	2.500	10.0	1.0	1.0	0.215	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	5.000	10.0	1.0	1.0	0.399	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	10.000	10.0	1.0	1.0	0.798	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	12.500	10.0	1.0	1.0	1.087	STD7	2025-10-09 12:55:27 UTC-4	used

Sulfate (Anions) $(\mu\text{S}/\text{cm}) \times \text{min}$ Function: $A = -0.0307148 + 7.61403E-3 \times Q$

Relative standard deviation 2.830893 %

Correlation coefficient 0.999580


Sample type	Index	Conc.	Volume	Dilution	Sample amount	Area	Ident	Date	Used
Standard 1	1	0.000	10.0	1.0	1.0	n. d.	STD1	2025-10-09 10:46:54 UTC-4	used
Standard 2	1	3.000	10.0	1.0	1.0	0.220	STD2	2025-10-09 11:08:18 UTC-4	used
Standard 3	1	6.000	10.0	1.0	1.0	0.433	STD3	2025-10-09 11:29:42 UTC-4	used
Standard 4	1	7.500	10.0	1.0	1.0	0.541	STD4	2025-10-09 11:51:08 UTC-4	used
Standard 5	1	15.000	10.0	1.0	1.0	1.078	STD5	2025-10-09 12:12:34 UTC-4	used
Standard 6	1	30.000	10.0	1.0	1.0	2.216	STD6	2025-10-09 12:34:00 UTC-4	used
Standard 7	1	37.000	10.0	1.0	1.0	2.827	STD7	2025-10-09 12:55:27 UTC-4	used

Sample data

Ident STD1
Sample type Standard 1
Determination start 2025-10-09 10:46:54 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.57 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions


Sample data

Ident STD2
Sample type Standard 2
Determination start 2025-10-09 11:08:18 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.34 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.392	0.0605	0.366	0.440	Fluoride
2	6.680	0.0597	0.264	0.646	Chloride
3	8.865	0.1310	0.444	0.652	Nitrite
4	11.008	0.0868	0.245	2.141	Bromide
5	12.692	0.1222	0.301	0.541	Nitrate
6	14.212	0.0833	0.158	1.118	Phosphate
7	15.883	0.2204	0.426	3.299	Sulfate


Sample data

Ident STD3
Sample type Standard 3
Determination start 2025-10-09 11:29:42 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.63 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.388	0.1150	0.718	0.813	Fluoride
2	6.662	0.1185	0.531	1.225	Chloride
3	8.833	0.2598	0.891	1.222	Nitrite
4	10.962	0.1713	0.487	4.073	Bromide
5	12.637	0.2429	0.605	1.019	Nitrate
6	14.187	0.1538	0.307	1.947	Phosphate
7	15.888	0.4330	0.848	6.091	Sulfate


Sample data

Ident STD4
Sample type Standard 4
Determination start 2025-10-09 11:51:08 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.34 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.392	0.1414	0.897	0.993	Fluoride
2	6.655	0.1461	0.654	1.497	Chloride
3	8.818	0.3220	1.103	1.496	Nitrite
4	10.933	0.2116	0.604	4.994	Bromide
5	12.607	0.3009	0.750	1.249	Nitrate
6	14.185	0.2152	0.426	2.669	Phosphate
7	15.900	0.5409	1.049	7.507	Sulfate


Sample data

Ident STD5
Sample type Standard 5
Determination start 2025-10-09 12:12:34 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.06 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.390	0.2819	1.816	1.954	Fluoride
2	6.645	0.2909	1.308	2.923	Chloride
3	8.803	0.6446	2.221	2.922	Nitrite
4	10.910	0.4207	1.208	9.773	Bromide
5	12.582	0.6001	1.504	2.435	Nitrate
6	14.178	0.3988	0.783	4.828	Phosphate
7	15.905	1.0776	2.103	14.557	Sulfate


Sample data

Ident STD6
Sample type Standard 6
Determination start 2025-10-09 12:34:00 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.06 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.388	0.5707	3.761	3.928	Fluoride
2	6.638	0.5955	2.722	5.923	Chloride
3	8.797	1.3228	4.622	5.920	Nitrite
4	10.897	0.8574	2.493	19.752	Bromide
5	12.565	1.2307	3.128	4.934	Nitrate
6	14.170	0.7975	1.608	9.517	Phosphate
7	15.900	2.2162	4.397	29.510	Sulfate


Sample data

Ident STD7
Sample type Standard 7
Determination start 2025-10-09 12:55:27 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.01 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.395	0.7379	4.813	5.072	Fluoride
2	6.643	0.7642	3.464	7.585	Chloride
3	8.805	1.7004	5.904	7.589	Nitrite
4	10.905	1.0987	3.182	25.268	Bromide
5	12.572	1.5811	4.005	6.323	Nitrate
6	14.172	1.0869	2.232	12.920	Phosphate
7	15.905	2.8273	5.621	37.537	Sulfate


Sample data

Ident ICV
Sample type Check standard 1
Determination start 2025-10-09 13:16:55 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.23 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.383	0.2843	1.850	1.970	Fluoride
2	6.632	0.2904	1.314	2.918	Chloride
3	8.785	0.6491	2.249	2.942	Nitrite
4	10.887	0.4238	1.222	9.842	Bromide
5	12.553	0.5909	1.487	2.398	Nitrate
6	14.152	0.4137	0.815	5.003	Phosphate
7	15.880	1.0887	2.130	14.702	Sulfate

Sample data

Ident ICB
Sample type Sample
Determination start 2025-10-09 13:38:24 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.12 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions


Sample data

Ident CCV
Sample type Check standard 1
Determination start 2025-10-15 09:11:15 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.40 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.402	0.2941	1.868	2.037	Fluoride
2	6.708	0.3004	1.339	3.017	Chloride
3	8.905	0.6680	2.271	3.025	Nitrite
4	11.060	0.4378	1.236	10.163	Bromide
5	12.747	0.6058	1.503	2.457	Nitrate
6	13.973	0.4015	0.808	4.859	Phosphate
7	15.970	1.1119	2.165	15.007	Sulfate

Sample data

Ident CCB
Sample type Sample
Determination start 2025-10-15 09:32:48 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.23 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions


Sample data

Ident LB137519BLS
Sample type Sample
Determination start 2025-10-15 09:54:20 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.06 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions




Sample data

Ident LB137519BSS
Sample type Check standard 1
Determination start 2025-10-15 10:15:51 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.12 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.400	0.2943	1.900	2.039	Fluoride
2	6.672	0.3036	1.373	3.048	Chloride
3	8.843	0.6686	2.311	3.028	Nitrite
4	10.963	0.4393	1.263	10.198	Bromide
5	12.642	0.6121	1.539	2.482	Nitrate
6	13.958	0.4303	0.874	5.199	Phosphate
7	15.992	1.1215	2.198	15.132	Sulfate


Sample data

Ident Q3341-01
Sample type Sample
Determination start 2025-10-15 10:37:24 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.01 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.432	0.0421	0.242	0.314	Fluoride
2	6.703	3.0073	14.272	29.678	Chloride
3	12.660	0.0072	0.017	0.085	Nitrate
4	14.012	0.0105	0.020	0.262	Phosphate
5	16.023	0.0522	0.102	1.089	Sulfate


Sample data

Ident Q3341-01MS
Sample type Sample
Determination start 2025-10-15 10:58:57 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.01 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.402	0.3005	1.921	2.081	Fluoride
2	6.672	3.2185	15.141	31.758	Chloride
3	8.832	0.6677	2.307	3.024	Nitrite
4	10.943	0.4364	1.254	10.131	Bromide
5	12.618	0.6144	1.545	2.491	Nitrate
6	13.960	0.3524	0.711	4.283	Phosphate
7	16.000	1.1520	2.256	15.533	Sulfate


Sample data

Ident Q3341-01MSD
Sample type Sample
Determination start 2025-10-15 11:20:31 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.95 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.398	0.2991	1.931	2.071	Fluoride
2	6.667	3.2148	15.156	31.722	Chloride
3	8.825	0.6671	2.311	3.022	Nitrite
4	10.937	0.4362	1.256	10.127	Bromide
5	12.612	0.6149	1.548	2.493	Nitrate
6	13.957	0.3524	0.711	4.282	Phosphate
7	15.998	1.1517	2.257	15.529	Sulfate


Sample data

Ident Q3341-02
Sample type Sample
Determination start 2025-10-15 11:42:05 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.06 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area ($\mu\text{S/cm}$) x min	Height $\mu\text{S/cm}$	Concentration ppm	Component name
1	3.393	0.0247	0.154	0.195	Fluoride
2	6.657	0.1789	0.815	1.820	Chloride
3	15.992	0.0333	0.065	0.840	Sulfate


Sample data

Ident Q3341-03
Sample type Sample
Determination start 2025-10-15 12:03:40 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.06 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area ($\mu\text{S/cm}$) x min	Height $\mu\text{S/cm}$	Concentration ppm	Component name
1	3.398	0.0093	0.055	0.090	Fluoride
2	6.663	0.1515	0.683	1.550	Chloride
3	13.968	0.0249	0.049	0.431	Phosphate
4	15.993	0.0615	0.118	1.211	Sulfate


Sample data

Ident Q3341-01DLX10
Sample type Sample
Determination start 2025-10-15 12:25:15 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.18 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.397	0.0059	0.027	0.067	Fluoride
2	6.658	0.2669	1.207	2.686	Chloride
3	15.985	0.0069	0.013	0.494	Sulfate


Sample data

Ident CCV
Sample type Check standard 1
Determination start 2025-10-15 12:46:37 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 13.06 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions



Peak number	Retention time min	Area (µS/cm) x min	Height µS/cm	Concentration ppm	Component name
1	3.398	0.2932	1.893	2.031	Fluoride
2	6.657	0.3008	1.372	3.020	Chloride
3	8.818	0.6681	2.322	3.026	Nitrite
4	10.927	0.4383	1.265	10.175	Bromide
5	12.598	0.6099	1.541	2.474	Nitrate
6	13.945	0.4097	0.836	4.956	Phosphate
7	16.002	1.1148	2.196	15.045	Sulfate

Sample data

Ident CCB
Sample type Sample
Determination start 2025-10-15 13:08:07 UTC-4
Method IC1-100925
Operator

Anions

Data source Conductivity detector 1 (Eco IC 1)
Channel Conductivity
Recording time 19.0 min
Integration Automatically
Column type Metrosep A Supp 19 - 150/4.0
Eluent composition not defined
Flow 0.700 mL/min
Maximum flow monitored yes
Pressure 12.95 MPa
Maximum pressure monitored yes
Temperature ---- °C

Anions