

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB137527

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/15/2025	08:20
2	CAL2	1	Water	NA	NA	20.2	7.00	10/15/2025	08:21
3	CAL3	1	Water	NA	NA	20.2	10.02	10/15/2025	08:22
4	ICV	1	Water	NA	NA	20.3	7.01	10/15/2025	08:25
5	CCV1	1	Water	NA	NA	20.3	2.01	10/15/2025	08:30
6	Q3345-07	1	Water	NA	NA	20.7	6.89	10/15/2025	08:45
7	Q3345-07DUP	1	Water	NA	NA	20.9	6.90	10/15/2025	08:47
8	CCV2	1	Water	NA	NA	20.3	12.02	10/15/2025	08:50