



Analytical Summary Report

Analysis Method: SM2120 B

Parameter: Color

Run Number: LB137535

Reviewed By: Iwona

Supervisor Review By: jignesh

pH Meter ID: WC pH Meter-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP115190
Calibration std, 5 color unit	WP115191
Calibration std, 10 color unit	WP115192
Calibration std, 15 color unit	WP115193
Calibration std, 20 color unit	WP115194
Calibration std, 25 color unit	WP115195
Calibration std, 30 color unit	WP115196
Calibration std, 40 color unit	WP115197
Calibration std, 50 color unit	WP115198
Calibration std, 60 color unit	WP115199
Calibration std, 70 color unit	WP115200
Calibration std, 100 color unit	WP115201

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB137535BL	LB137535BL	1	5.02	N/A	0.00	10/15/2025	10:30
2	Q3344-01	1 EFFLUENT-COMPOSITE	20	7.98	N/A	40.00	10/15/2025	10:35
3	Q3344-01DUP	1 EFFLUENT-COMPOSITE	20	7.97	N/A	40.00	10/15/2025	10:40