

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137538

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/15/2025	10:15
2	CAL2	1	Water	NA	NA	20.2	7.01	10/15/2025	10:16
3	CAL3	1	Water	NA	NA	20.3	10.02	10/15/2025	10:17
4	ICV	1	Water	NA	NA	20.2	7.00	10/15/2025	10:19
5	CCV1	1	Water	NA	NA	20.3	2.02	10/15/2025	10:22
6	Q3342-02	1	Solid	20.02	20	20.5	10.97	10/15/2025	10:30
7	Q3342-02DUP	1	Solid	20.03	20	20.6	10.99	10/15/2025	10:33
8	Q3343-04	1	Solid	20.02	20	20.1	9.37	10/15/2025	10:40
9	Q3345-02	1	Solid	20.03	20	20.3	8.92	10/15/2025	10:50
10	Q3345-04	1	Solid	20.04	20	20.2	10.05	10/15/2025	11:00
11	Q3345-06	1	Solid	20.03	20	20.1	9.73	10/15/2025	11:10
12	Q3348-04	1	Solid	20.02	20	20.7	5.56	10/15/2025	11:20
13	Q3348-08	1	Solid	20.03	20	20.5	5.96	10/15/2025	11:30
14	CCV2	1	Water	NA	NA	20.3	12.02	10/15/2025	11:33