

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137551

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/16/2025	07:50
2	CAL2	1	Water	NA	NA	20.2	7.00	10/16/2025	07:51
3	CAL3	1	Water	NA	NA	20.3	10.02	10/16/2025	07:55
4	ICV	1	Water	NA	NA	20.2	6.98	10/16/2025	08:00
5	CCV1	1	Water	NA	NA	20.2	2.01	10/16/2025	08:02
6	Q3328-01	1	Water	NA	NA	20.2	4.88	10/16/2025	08:07
7	Q3328-01DUP	1	Water	NA	NA	20.3	4.90	10/16/2025	08:10
8	Q3356-01	1	Water	NA	NA	20.6	7.20	10/16/2025	08:20
9	Q3356-01DUP	1	Water	NA	NA	20.8	7.22	10/16/2025	08:22
10	CCV2	1	Water	NA	NA	20.3	12.02	10/16/2025	08:25

WORKLIST(Hardcopy Internal Chain)

W3137551

WorkList Name : ph w q3356 WorkList ID : 192487 Department : Wet-Chemistry Date : 10-16-2025 07:31:34

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3328-01	A MH-10-10-25	Water	pH	Cool 4 deg C	EURO03	D31	10/10/2025	9040C
Q3356-01	D AUD-25-0169-AUD-25-0170	Water	pH	Cool 4 deg C	PSEG03	D31	10/15/2025	9040C

Date/Time 10/16/25 07:45
 Raw Sample Received by: JP Wey
 Raw Sample Relinquished by: JP Wey

Date/Time 10/16/25 12:10
 Raw Sample Received by: JP Wey
 Raw Sample Relinquished by: JP Wey