

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137563

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	10/16/2025	11:45
2	Q3356-01		1	13	100.00	>212.0	>212.0	10/16/2025	12:15
3	Q3356-01DUP		1	13	100.00	>212.0	>212.0	10/16/2025	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66137563

WorkList Name : fp-1016

WorkList ID : 192503

Department : Wet-Chemistry

Date : 10-16-2025 11:13:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3356-01	AUD-25-0169-AUD-25-0170	Water	Flash Point	Cool 4 deg C	PSEG03	D31	10/15/2025	1010B

Date/Time 10/16/2025 11:35
Raw Sample Received by: RM WWS
Raw Sample Relinquished by: JRC WWS

Date/Time 10/16/2025 13:25
Raw Sample Received by: JRC WWS
Raw Sample Relinquished by: JRC WWS