

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137571

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	10/17/2025	08:35
2	Q3358-01		1	11	100.00	>212.0	>212.0	10/17/2025	09:05
3	Q3358-01DUP		1	11	100.00	>212.0	>212.0	10/17/2025	09:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137571

WorkList Name : fp-10-16=

WorkList ID : 192516

Department : Wet-Chemistry

Date : 10-16-2025 16:26:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3358-01	NWB-2018	Solid	Flash Point	Cool 4 deg C	PSEG03	D31	10/15/2025	1010B

Date/Time 10/17/2025 08:20
 Raw Sample Received by: RM WWC
 Raw Sample Relinquished by: RM WWC

Date/Time 10/17/2025 10:15
 Raw Sample Received by: RM WWC
 Raw Sample Relinquished by: RM WWC