



PERCENT SOLID

Supervisor: Iwona  
Analyst: rubina  
Date: 10/20/2025

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 10/17/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 09:25  
Out Date: 10/18/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB137576

| Lab ID   | Client SampleID            | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments |
|----------|----------------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|----------|
| Q3373-05 | SVOC-GPC-BLANK             | 1      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |          |
| Q3373-06 | PEST-GPC-BLANK             | 2      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |          |
| Q3373-07 | PEST-GPC-BLANK-SPIKE       | 3      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |          |
| Q3373-08 | SVOC-GPC2-BLANK            | 4      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |          |
| Q3373-09 | PEST-GPC2-BLANK            | 5      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |          |
| Q3373-10 | PEST -GPC2-BLANK-SPIKE     | 6      | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   |          |
| Q3374-01 | TP-4-WC                    | 7      | 1.15           | 10.69        | 11.84                   | 10.25                     | 85.1    |          |
| Q3374-02 | TP-4-WC-EPH                | 8      | 1.11           | 10.86        | 11.97                   | 11.38                     | 94.6    |          |
| Q3374-03 | TP-4-WC-VOC                | 9      | 1.15           | 10.45        | 11.6                    | 10.31                     | 87.7    |          |
| Q3374-05 | TP-5-WC                    | 10     | 1.13           | 10.30        | 11.43                   | 10.72                     | 93.1    |          |
| Q3374-06 | TP-5-WC-EPH                | 11     | 1.17           | 10.34        | 11.51                   | 9.95                      | 84.9    |          |
| Q3374-07 | TP-5-WC-VOC                | 12     | 1.18           | 10.94        | 12.12                   | 11.19                     | 91.5    |          |
| Q3375-01 | PR132-S02-074099-20251016  | 13     | 1.12           | 10.21        | 11.33                   | 8.19                      | 69.2    |          |
| Q3375-02 | PR132-S04-000013-20251016  | 14     | 1.15           | 10.84        | 11.99                   | 9.25                      | 74.7    |          |
| Q3375-03 | PR132-S05-000057-20251016  | 15     | 1.16           | 10.58        | 11.74                   | 8.22                      | 66.7    |          |
| Q3375-04 | PR132-S02-010074-20251016  | 16     | 1.14           | 10.34        | 11.48                   | 6.86                      | 55.3    |          |
| Q3375-05 | PR132-DP02-074099-20251016 | 17     | 1.15           | 10.84        | 11.99                   | 8.38                      | 66.7    |          |
| Q3376-01 | PR132-S05-057082-20251016  | 18     | 1.13           | 10.84        | 11.97                   | 6.93                      | 53.5    |          |
| Q3376-02 | PR132-DP01-20251016        | 19     | 1.16           | 10.07        | 11.23                   | 6.56                      | 53.6    |          |
| Q3376-03 | PR132-COMP2-20251016       | 20     | 1.11           | 10.13        | 11.24                   | 6.45                      | 52.7    |          |
| Q3376-04 | PR132-S12-047072-20251016  | 21     | 1.15           | 10.70        | 11.85                   | 6.64                      | 51.3    |          |
| Q3376-05 | PR132-S12-020047-20251016  | 22     | 1.13           | 10.49        | 11.62                   | 8.14                      | 66.8    |          |
| Q3381-01 | VNJ-222                    | 23     | 1.18           | 10.38        | 11.56                   | 10.46                     | 89.4    |          |
| Q3381-03 | VNJ-245                    | 24     | 1.18           | 10.64        | 11.82                   | 9.86                      | 81.6    |          |
| Q3381-05 | RBR200059                  | 25     | 1.11           | 10.71        | 11.82                   | 10.6                      | 88.6    |          |



# PERCENT SOLID

Supervisor: Iwona  
Analyst: rubina  
Date: 10/20/2025

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 10/17/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 09:25  
Out Date: 10/18/2025  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: % SOLID- OVEN

QC:LB137576

| Lab ID   | Client SampleID      | Dish # | Dish Wt(g) (A) | Sample Wt(g) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments        |
|----------|----------------------|--------|----------------|--------------|-------------------------|---------------------------|---------|-----------------|
| Q3381-07 | VNJ-255              | 26     | 1.16           | 10.11        | 11.27                   | 10.51                     | 92.5    |                 |
| Q3382-01 | CENTRAL-AVE-TP       | 27     | 1.14           | 10.73        | 11.87                   | 10.25                     | 84.9    |                 |
| Q3382-02 | CENTRAL-AVE-TP-EPH-2 | 28     | 1.16           | 10.81        | 11.97                   | 9.78                      | 79.7    |                 |
| Q3383-01 | OR-03-101725         | 29     | 1.13           | 10.60        | 11.73                   | 10.88                     | 92.0    |                 |
| Q3383-02 | OR-03-101725-E2      | 30     | 1.14           | 10.91        | 12.05                   | 11.07                     | 91.0    |                 |
| Q3386-01 | 237                  | 38     | 1.15           | 10.68        | 11.83                   | 11.2                      | 94.1    |                 |
| Q3386-02 | 237-E2               | 39     | 1.15           | 10.85        | 12.00                   | 11.2                      | 92.6    |                 |
| Q3387-01 | BC289050-1-1         | 40     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC            |
| Q3387-02 | BC289050-1-2         | 41     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC            |
| Q3387-03 | BC289050-2-1         | 42     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC            |
| Q3387-04 | BC289050-2-2         | 43     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC            |
| Q3388-01 | 461                  | 44     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | soily-debris    |
| Q3389-01 | KMH472J-1-1          | 45     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC            |
| Q3389-02 | KMH472J-1-2          | 46     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | PILC            |
| Q3390-01 | 4175                 | 47     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | debris          |
| Q3390-03 | 340                  | 48     | 1.18           | 10.76        | 11.94                   | 10.34                     | 85.1    |                 |
| Q3390-04 | 340-E2               | 49     | 1.16           | 10.88        | 12.04                   | 9.11                      | 73.1    |                 |
| Q3391-01 | RCA                  | 31     | 1.00           | 1.00         | 2.00                    | 2.00                      | 100.0   | Concrete sample |
| Q3392-01 | MH-20                | 32     | 1.18           | 10.77        | 11.95                   | 10.9                      | 90.3    |                 |
| Q3392-02 | MH-20-EPH            | 33     | 1.16           | 11.00        | 12.16                   | 11.07                     | 90.1    |                 |
| Q3392-03 | MH-20-VOC            | 34     | 1.18           | 10.89        | 12.07                   | 10.69                     | 87.3    |                 |
| Q3392-05 | MH-19                | 35     | 1.16           | 10.50        | 11.66                   | 10.34                     | 87.4    |                 |
| Q3392-06 | MH-19-EPH            | 36     | 1.12           | 10.87        | 11.99                   | 10.41                     | 85.5    |                 |
| Q3392-07 | MH-19-VOC            | 37     | 1.16           | 10.50        | 11.66                   | 10.37                     | 87.7    |                 |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$