

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137579

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	28.00	82.4	82.7	10/20/2025	08:25
2	Q3390-05		1	15	100.00	>212.0	>212.0	10/20/2025	08:55
3	Q3390-05DUP		1	15	100.00	>212.0	>212.0	10/20/2025	09:25
4	Q3390-06		1	13	100.00	>212.0	>212.0	10/20/2025	09:55
5	Q3390-07		1	15	100.00	>212.0	>212.0	10/20/2025	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137579

WorkList Name : fp-10-17.

WorkList ID : 192538

Department : Wet-Chemistry

Date : 10-17-2025 15:04:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3390-05	4174	Water	Flash Point	Cool 4 deg C	PSEG03	D41	10/17/2025	1010B
Q3390-06	4160	Water	Flash Point	Cool 4 deg C	PSEG03	D41	10/17/2025	1010B
Q3390-07	4090	Water	Flash Point	Cool 4 deg C	PSEG03	D41	10/17/2025	1010B

Date/Time 10/17/2025 08.10
 Raw Sample Received by: RM (w/c)
 Raw Sample Relinquished by: RM (w/c)

Date/Time 10/17/2025 11.16
 Raw Sample Received by: RM (w/c)
 Raw Sample Relinquished by: RM (w/c)