

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB137580

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/17/2025	16:25
2	CAL2	1	Water	NA	NA	20.2	7.00	10/17/2025	16:26
3	CAL3	1	Water	NA	NA	20.2	10.02	10/17/2025	16:29
4	ICV	1	Water	NA	NA	20.3	7.00	10/17/2025	16:30
5	CCV1	1	Water	NA	NA	20.3	2.01	10/17/2025	16:33
6	Q3369-04	1	Solid	20.02	20	20.7	8.86	10/17/2025	16:35
7	Q3369-08	1	Solid	20.03	20	20.4	8.30	10/17/2025	16:37
8	Q3374-04	1	Solid	20.02	20	20.9	6.84	10/17/2025	16:40
9	Q3374-08	1	Solid	20.03	20	20.1	6.75	10/17/2025	16:45
10	Q3381-02	1	Solid	20.02	20	20.6	5.31	10/17/2025	16:55
11	Q3381-04	1	Solid	20.03	20	20.1	5.91	10/17/2025	17:00
12	Q3381-06	1	Solid	20.02	20	20.2	6.03	10/17/2025	17:10
13	Q3381-08	1	Solid	20.03	20	20.1	6.30	10/17/2025	17:15
14	Q3388-02	1	Solid	20.02	20	20.7	6.70	10/17/2025	17:25
15	Q3390-02	1	Solid	20.01	20	20.7	6.70	10/17/2025	17:30
16	CCV2	1	Water	NA	NA	20.3	12.02	10/17/2025	17:35
17	Q3392-04	1	Solid	20.02	20	20.4	5.76	10/17/2025	17:35
18	Q3392-08	1	Solid	20.03	20	20.2	6.37	10/17/2025	17:40
19	Q3392-08DUP	1	Solid	20.04	20	20.2	6.39	10/17/2025	17:44
20	CCV3	1	Water	NA	NA	20.2	2.01	10/17/2025	17:48

WORKLIST(Hardcopy Internal Chain)

137580

WorkList Name : corrsivity q3374 WorkList ID : 192526 Department : Wet-Chemistry Date : 10-17-2025 10:48:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3369-04	MH-22	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/16/2025	9045D
Q3369-08	MH-21	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/16/2025	9045D
Q3374-04	TP-4-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	10/16/2025	9045D
Q3374-08	TP-5-WC	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	10/17/2025	9045D
Q3381-02	VNJ-222	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/17/2025	9045D
Q3381-04	VNJ-245	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/17/2025	9045D
Q3381-06	RBR200059	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/17/2025	9045D
Q3381-08	VNJ-255	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/17/2025	9045D
Q3388-02	461	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/17/2025	9045D
Q3390-02	4175	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/17/2025	9045D
Q3392-04	MH-20	Solid	Corrosivity	Cool 4 deg C	PSEG03	D41	10/17/2025	9045D
Q3392-08	MH-19	Solid	Corrosivity	Cool 4 deg C	PSEG03	D31	10/17/2025	9045D

Date/Time 10/17/25 16:20

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 10/17/25 19:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]