

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB137581

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/17/2025	16:45
2	CAL2	1	Water	NA	NA	20.1	7.00	10/17/2025	16:46
3	CAL3	1	Water	NA	NA	20.2	10.02	10/17/2025	16:48
4	ICV	1	Water	NA	NA	20.2	6.99	10/17/2025	16:50
5	CCV1	1	Water	NA	NA	20.2	2.01	10/17/2025	16:55
6	Q3386-03	1	Water	NA	NA	20.1	6.50	10/17/2025	17:00
7	Q3386-03DUP	1	Water	NA	NA	20.2	6.52	10/17/2025	17:05
8	Q3386-04	1	Water	NA	NA	20.1	6.60	10/17/2025	17:15
9	Q3386-04DUP	1	Water	NA	NA	20.3	6.63	10/17/2025	17:19
10	Q3390-05	1	Water	NA	NA	20.1	7.57	10/17/2025	17:25
11	Q3390-05DUP	1	Water	NA	NA	20.3	7.60	10/17/2025	17:30
12	Q3390-06	1	Water	NA	NA	20.2	7.11	10/17/2025	17:35
13	Q3390-06DUP	1	Water	NA	NA	20.2	7.12	10/17/2025	17:36
14	Q3390-07	1	Water	NA	NA	20.2	7.07	10/17/2025	17:40
15	Q3390-07DUP	1	Water	NA	NA	20.4	7.09	10/17/2025	17:44
16	CCV2	1	Water	NA	NA	20.3	12.02	10/17/2025	17:45

WORKLIST(Hardcopy Internal Chain)

134581

WorkList Name : ph q 3390 WorkList ID : 192537 Department : Wet-Chemistry Date : 10-17-2025 16:28:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3386-03	C 3443	Water	pH	Cool 4 deg C	PSEG03	D41	10/17/2025	9040C
Q3386-04	C 3444	Water	pH	Cool 4 deg C	PSEG03	D41	10/17/2025	9040C
Q3390-05	C 4174	Water	pH	Cool 4 deg C	PSEG03	D41	10/17/2025	9040C
Q3390-06	C 4160	Water	pH	Cool 4 deg C	PSEG03	D41	10/17/2025	9040C
Q3390-07	C 4090	Water	pH	Cool 4 deg C	PSEG03	D41	10/17/2025	9040C

Date/Time 10/17/25 16:35
 Raw Sample Received by: JH CEC
 Raw Sample Relinquished by: CACSM1

Date/Time 10/17/25 18:30
 Raw Sample Received by: CACSM1
 Raw Sample Relinquished by: JH CEC