

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137606

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/22/2025	08:35
2	CAL2	1	Water	NA	NA	20.2	7.00	10/22/2025	08:36
3	CAL3	1	Water	NA	NA	20.2	10.02	10/22/2025	08:37
4	ICV	1	Water	NA	NA	20.3	7.01	10/22/2025	08:40
5	CCV1	1	Water	NA	NA	20.2	2.02	10/22/2025	08:44
6	Q3398-01	1	Water	NA	NA	20.2	9.43	10/22/2025	09:00
7	Q3398-01DUP	1	Water	NA	NA	20.3	9.45	10/22/2025	09:02
8	Q3400-01	1	Water	NA	NA	20.1	7.62	10/22/2025	09:10
9	Q3400-01DUP	1	Water	NA	NA	20.2	7.64	10/22/2025	09:11
10	Q3417-01	1	Water	NA	NA	20.3	5.60	10/22/2025	09:15
11	Q3417-01DUP	1	Water	NA	NA	20.4	5.62	10/22/2025	09:18
12	CCV2	1	Water	NA	NA	20.3	12.02	10/22/2025	09:20

WORKLIST(Hardcopy Internal Chain)

137606

WorkList Name : ph w q3417

WorkList ID : 192595

Department : Wet-Chemistry

Date : 10-22-2025 08:14:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3398-01	A748	Water	pH	Cool 4 deg C	PSEG03	D31	10/20/2025	9040C
Q3400-01	FRAC-TANK-0760450	Water	pH	Cool 4 deg C	ICES02	D31	10/20/2025	9040C
Q3417-01	101425	Water	pH	Cool 4 deg C	PSEG03	D31	10/21/2025	9040C

Date/Time 10/22/25 08:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10/22/25 12:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]