

LB137620


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

40287



A= 0.0000e+000

B= 2.4697e-004

C= 1.9379e-002

Rho= 0.9999695

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	-22	0.000	-22	0			
0.2	0.200	0.192	-0.008	697	0.0 %	697	0			
2.5	2.500	2.460	-0.040	9881	0.0 %	9881	0			
5.0	5.000	5.035	0.035	20307	0.0 %	20307				
7.5	7.500	7.531	0.031	30414	0.0 %	30414				
10.0	10.000	9.969	-0.031	40287	0.0 %	40287				

LB137620 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Oct 2025 07:49:44

Type	Sample ID	Extended ID	Conc.	µ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	-22	PPB	Std	0.0000	7470A	23 Oct 2025 08:00:59
S	0.2 - 1	51.2	-	697	PPB	Std	0.2000	7470A	23 Oct 2025 08:03:23
S	2.5 - 1	52.5	-	9881	PPB	Std	2.5000	7470A	23 Oct 2025 08:06:40
S	5.0 - 1	55	-	20307	PPB	Std	5.0000	7470A	23 Oct 2025 08:07:57
S	7.5 - 1	57.8	-	30414	PPB	Std	7.5000	7470A	23 Oct 2025 08:10:16
S	10.0 - 1	58.8	-	40287	PPB	Std	10.0000	7470A	23 Oct 2025 08:12:35
U	ICV84 - 1	ICV84	3.9860	16061	PPB	SMPL	-7470A		23 Oct 2025 08:15:49
U	ICB84 - 1	ICB84	-0.0219	-167	PPB	SMPL	-7470A		23 Oct 2025 08:18:04
U	CCV75 - 1	CCV75	5.0337	20303	PPB	SMPL	-7470A		23 Oct 2025 08:22:44
U	CCB75 - 1	CCB75	-0.0283	-193	PPB	SMPL	-7470A		23 Oct 2025 08:24:59
U	CRA - 1	CRA	0.2207	815	PPB	SMPL	-7470A		23 Oct 2025 08:27:18
U	HighStd - 1	HighStd	10.1090	40853	PPB	SMPL	-7470A		23 Oct 2025 08:29:33
U	ChkStd - 1	ChkStd	6.7791	27370	PPB	SMPL	-7470A		23 Oct 2025 08:31:48
U	PB170219BL - 1	PBW	-0.0515	-287	PPB	SMPL	-7470A		23 Oct 2025 08:34:13
U	PB170219BS - 1	LCSW	3.7391	15061	PPB	SMPL	-7470A		23 Oct 2025 08:36:34
U	Q3368-06 - 1	OUTSIDE-DUMPSTER	-0.0214	-165	PPB	SMPL	-7470A		23 Oct 2025 08:38:51
U	Q3368-07 - 1	PRESS-SLUDGE	0.0236	17	PPB	SMPL	-7470A		23 Oct 2025 08:41:10
U	Q3397-04 - 1	TP-7-WC	0.0283	36	PPB	SMPL	-7470A		23 Oct 2025 08:43:26
U	Q3399-02 - 1	COMP-ROLLOFF	0.0236	17	PPB	SMPL	-7470A		23 Oct 2025 08:45:43
U	Q3402-03 - 1	41-LA	0.0258	26	PPB	SMPL	-7470A		23 Oct 2025 08:48:06
U	CCV76 - 1	CCV76	5.0485	20363	PPB	SMPL	-7470A		23 Oct 2025 08:50:21
U	CCB76 - 1	CCB76	-0.0219	-167	PPB	SMPL	-7470A		23 Oct 2025 08:52:36
U	Q3402-06 - 1	41-RA	0.0248	22	PPB	SMPL	-7470A		23 Oct 2025 08:54:55
U	Q3402-09 - 1	41-LB	0.0307	46	PPB	SMPL	-7470A		23 Oct 2025 08:57:11
U	Q3402-12 - 1	41-RB	0.0367	70	PPB	SMPL	-7470A		23 Oct 2025 08:59:26
U	Q3403-04 - 1	MH-18	0.0298	42	PPB	SMPL	-7470A		23 Oct 2025 09:01:43
U	Q3404-02 - 1	MOO-25-0285-0288	0.0248	22	PPB	SMPL	-7470A		23 Oct 2025 09:04:06
U	Q3404-06 - 1	MOO-25-0292	0.0248	22	PPB	SMPL	-7470A		23 Oct 2025 09:06:22
U	Q3406-02 - 1	ARS20-0037	0.0115	-32	PPB	SMPL	-7470A		23 Oct 2025 09:08:38
U	Q3412-04 - 1	TP-6	0.0181	-5	PPB	SMPL	-7470A		23 Oct 2025 09:10:55
U	Q3412-04DUP - 1	TP-6DUP	0.0246	21	PPB	SMPL	-7470A		23 Oct 2025 09:13:42
U	Q3412-04MS - 1	TP-6MS	3.8216	15395	PPB	SMPL	-7470A		23 Oct 2025 09:16:02
U	CCV77 - 1	CCV77	4.9033	19775	PPB	SMPL	-7470A		23 Oct 2025 09:18:17
U	CCB77 - 1	CCB77	-0.0216	-166	PPB	SMPL	-7470A		23 Oct 2025 09:20:36
U	Q3412-04MSD - 1	TP-6MSD	3.8082	15341	PPB	SMPL	-7470A		23 Oct 2025 09:22:54
U	PB170172TB - 1	PB170172TB	-0.0006	-81	PPB	SMPL	-7470A		23 Oct 2025 09:25:09
U	Q3412-04LX5 - 1		0.0258	26	PPB	SMPL	-7470A		23 Oct 2025 09:27:36
U	Q3412-04A - 1		4.0633	16374	PPB	SMPL	-7470A		23 Oct 2025 09:33:49
U	CCV78 - 1	CCV78	5.0493	20366	PPB	SMPL	-7470A		23 Oct 2025 09:36:05
U	CCB78 - 1	CCB78	-0.0327	-211	PPB	SMPL	-7470A		23 Oct 2025 09:38:24