

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137667

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/28/2025	09:35
2	CAL2	1	Water	NA	NA	20.3	7.00	10/28/2025	09:36
3	CAL3	1	Water	NA	NA	20.2	10.02	10/28/2025	09:37
4	ICV	1	Water	NA	NA	20.2	7.00	10/28/2025	09:39
5	CCV1	1	Water	NA	NA	20.3	2.01	10/28/2025	09:40
6	Q3472-03	1	Water	NA	NA	20.1	5.04	10/28/2025	09:45
7	Q3472-03DUP	1	Water	NA	NA	20.2	5.05	10/28/2025	09:47
8	Q3472-19	1	Water	NA	NA	20.5	7.57	10/28/2025	09:55
9	Q3472-19DUP	1	Water	NA	NA	20.7	7.59	10/28/2025	10:00
10	Q3472-20	1	Water	NA	NA	20.1	6.19	10/28/2025	10:10
11	Q3472-20DUP	1	Water	NA	NA	20.2	6.20	10/28/2025	10:11
12	CCV2	1	Water	NA	NA	20.3	12.02	10/28/2025	10:15

WORKLIST(Hardcopy Internal Chain)

VB 137667

WorkList Name : PH W Q3466 WorkList ID : 192728 Department : Wet-Chemistry Date : 10-28-2025 09:19:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3472-03	COMP-7	Water	pH	Cool 4 deg C	PSEG03	D11	10/27/2025	9040C
Q3472-19	159	Water	pH	Cool 4 deg C	PSEG03	D41	10/27/2025	9040C
Q3472-20	COMP-6	Water	pH	Cool 4 deg C	PSEG03	D41	10/27/2025	9040C

Date/Time 10/28/25 09:25
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 10/28/25 12:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]