

LB137668


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

5370E



A= 0.0000e+000

B= 1.8787e-004

C= -3.7871e-002

Rho= 0.9997108

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.033	-0.033	26	0.000	26				
0.2	0.200	0.170	-0.030	1104	0.0 %	1104				
2.5	2.500	2.672	0.172	14424	0.0 %	14424				
5.0	5.000	4.915	-0.085	26364	0.0 %	26364				
7.5	7.500	7.424	-0.076	39720	0.0 %	39720				
10.0	10.000	10.052	0.052	53708	0.0 %	53708				

LB137668 INSTRUMENT ID: CV1

Method: 7471B		Operator: Admin		Date of Analysis: 28 Oct 2025 12:28:20					
Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
U	PB170271BS - 1	LCSS	3.9664	21374	PPB	SMPL	-7471B		28 Oct 2025 13:14:11
S	0.0 - 1	50	-	26	PPB	Std	0.0000	7471B	28 Oct 2025 12:32:55
S	0.2 - 1	102	-	1104	PPB	Std	0.2000	7471B	28 Oct 2025 12:35:13
S	2.5 - 1	525	-	14424	PPB	Std	2.5000	7471B	28 Oct 2025 12:37:30
S	5.0 - 1	55	-	26364	PPB	Std	5.0000	7471B	28 Oct 2025 12:39:46
S	7.5 - 1	55	-	39720	PPB	Std	7.5000	7471B	28 Oct 2025 12:42:22
S	10.0 - 1	55	-	53708	PPB	Std	10.0000	7471B	28 Oct 2025 12:45:04
U	ICV90 - 1	ICV90	4.2961	23069	PPB	SMPL	-7471B		28 Oct 2025 12:51:39
U	ICB90 - 1	ICB90	-0.0873	-263	PPB	SMPL	-7471B		28 Oct 2025 12:53:54
U	CCV97 - 1	CCV97	5.0411	27035	PPB	SMPL	-7471B		28 Oct 2025 12:56:34
U	CCB97 - 1	CCB97	-0.0931	-294	PPB	SMPL	-7471B		28 Oct 2025 12:58:49
U	CRA - 1	CRA	0.1620	1064	PPB	SMPL	-7471B		28 Oct 2025 13:04:08
U	HighStd - 1	HighStd	10.0580	53739	PPB	SMPL	-7471B		28 Oct 2025 13:06:22
U	ChkStd - 1	ChkStd	7.1100	38047	PPB	SMPL	-7471B		28 Oct 2025 13:08:38
U	PB170271BL - 1	PBS	-0.0933	-295	PPB	SMPL	-7471B		28 Oct 2025 13:11:27
U	Q3451-01 - 1	TP-11	0.0277	349	PPB	SMPL	-7471B		28 Oct 2025 13:16:28
U	Q3451-01DUP - 1	TP-11DUP	0.0260	340	PPB	SMPL	-7471B		28 Oct 2025 13:21:48
U	Q3451-01MS - 1	TP-11MS	3.9979	21482	PPB	SMPL	-7471B		28 Oct 2025 13:24:04
U	Q3451-01MSD - 1	TP-11MSD	4.1063	22059	PPB	SMPL	-7471B		28 Oct 2025 13:26:21
U	Q3451-05 - 1	TP-12	-0.0133	131	PPB	SMPL	-7471B		28 Oct 2025 13:29:02
U	CCV98 - 1	CCV98	4.7663	25572	PPB	SMPL	-7471B		28 Oct 2025 13:31:39
U	CCB98 - 1	CCB98	-0.0707	-175	PPB	SMPL	-7471B		28 Oct 2025 13:33:55
U	Q3452-01 - 1	JB-1	44.3148	236084	PPB	SMPL	-7471B		28 Oct 2025 13:36:34
U	Q3457-01 - 1	TR-05-102425	0.2054	1295	PPB	SMPL	-7471B		28 Oct 2025 13:38:50
U	Q3460-01 - 1	WC-1	3.6256	19500	PPB	SMPL	-7471B		28 Oct 2025 13:41:55
U	Q3460-05 - 1	WC-2	-0.0153	120	PPB	SMPL	-7471B		28 Oct 2025 13:44:12
U	Q3460-09 - 1	WC-3	0.0209	313	PPB	SMPL	-7471B		28 Oct 2025 13:46:51
U	Q3460-13 - 1	WC-4	4.4122	23687	PPB	SMPL	-7471B		28 Oct 2025 13:49:07
U	Q3460-17 - 1	WC-5	2.4777	13390	PPB	SMPL	-7471B		28 Oct 2025 13:51:24
U	Q3465-01 - 1	SU-03-10242025	0.2699	1638	PPB	SMPL	-7471B		28 Oct 2025 13:54:05
U	Q3466-02 - 1	B8-0.0-1.0-20251023	3.4824	18738	PPB	SMPL	-7471B		28 Oct 2025 13:56:39
U	Q3466-05 - 1	B9-0.0-1.0-20251023	0.5571	3167	PPB	SMPL	-7471B		28 Oct 2025 13:58:56
U	CCV99 - 1	CCV99	4.9932	26780	PPB	SMPL	-7471B		28 Oct 2025 14:01:32
U	CCB99 - 1	CCB99	-0.0961	-310	PPB	SMPL	-7471B		28 Oct 2025 14:03:50
U	Q3466-07 - 1	B7-1.0-2.0-20251023	0.2381	1469	PPB	SMPL	-7471B		28 Oct 2025 14:06:31
U	Q3469-01 - 1	TP-13	0.1631	1070	PPB	SMPL	-7471B		28 Oct 2025 14:08:47
U	Q3451-01LX5 - 1		-0.0407	-15	PPB	SMPL	-7471B		28 Oct 2025 14:11:02
U	Q3451-01A - 1		3.7734	20287	PPB	SMPL	-7471B		28 Oct 2025 14:13:18
U	Q3452-01DLX10 - 1		6.0962	32651	PPB	SMPL	-7471B		28 Oct 2025 14:20:48
U	CCV100 - 1	CCV100	4.9036	26303	PPB	SMPL	-7471B		28 Oct 2025 14:23:03
U	CCB100 - 1	CCB100	-0.0730	-187	PPB	SMPL	-7471B		28 Oct 2025 14:25:30