

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137674

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/28/2025	14:40
2	CAL2	1	Water	NA	NA	20.2	7.00	10/28/2025	14:41
3	CAL3	1	Water	NA	NA	20.3	10.02	10/28/2025	14:42
4	ICV	1	Water	NA	NA	20.3	7.00	10/28/2025	14:45
5	CCV1	1	Water	NA	NA	20.3	2.01	10/28/2025	14:50
6	Q3475-01	1	Water	NA	NA	20.1	6.13	10/28/2025	15:10
7	Q3475-01DUP	1	Water	NA	NA	20.3	6.15	10/28/2025	15:13
8	Q3477-03	1	Water	NA	NA	20.4	7.07	10/28/2025	15:20
9	Q3477-03DUP	1	Water	NA	NA	20.6	7.08	10/28/2025	15:23
10	CCV2	1	Water	NA	NA	20.3	12.02	10/28/2025	15:25

WORKLIST(Hardcopy Internal Chain)

UB 137674

WorkList Name : ph w q3475 WorkList ID : 192739 Department : Wet-Chemistry Date : 10-28-2025 14:11:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3475-01	B GW-DRUM	Water	pH	Cool 4 deg C	PSEG03	D41	10/28/2025	9040C
Q3477-03	F SOUTH-MAHWAH-WATER	Water	pH	Cool 4 deg C	PSEG03	D41	10/28/2025	9040C

Date/Time 10/28/25 14:30
 Raw Sample Received by: JO WWC
 Raw Sample Relinquished by: ITCsm

Date/Time 10/28/25 17:30
 Raw Sample Received by: ITCsm
 Raw Sample Relinquished by: JO WWC