

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137703

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/30/2025	11:10
2	CAL2	1	Water	NA	NA	20.2	7.00	10/30/2025	11:11
3	CAL3	1	Water	NA	NA	20.3	10.02	10/30/2025	11:14
4	ICV	1	Water	NA	NA	20.3	6.98	10/30/2025	11:15
5	CCV1	1	Water	NA	NA	20.2	2.01	10/30/2025	11:16
6	Q3484-01	1	Water	NA	NA	20.7	8.97	10/30/2025	11:35
7	Q3484-01DUP	1	Water	NA	NA	20.8	8.99	10/30/2025	11:37
8	Q3492-01	1	Water	NA	NA	20.4	7.53	10/30/2025	11:44
9	CCV2	1	Water	NA	NA	20.2	12.02	10/30/2025	11:45

WORKLIST(Hardcopy Internal Chain)

UR 137703

WorkList Name : ph q 3484

WorkList ID : 192794

Department : Wet-Chemistry

Date : 10-30-2025 10:57:02

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3484-01	TRE-25-0118	Water	pH	Cool 4 deg C	PSEG03	D41	10/29/2025	9040C
Q3492-01	252611	Water	pH	Cool 4 deg C	PSEG03	D31	10/29/2025	9040C

Date/Time 10/30/25 11:00
Raw Sample Received by: JQ WCL
Raw Sample Relinquished by: JQ WCL

Date/Time 10/30/25 13:35
Raw Sample Received by: JQ WCL
Raw Sample Relinquished by: JQ WCL