

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB137713

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3242

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	83.0	10/30/2025	09:30
2	Q3479-01		1	11	100.00	>212.0	>212.0	10/30/2025	10:30
3	Q3479-01DUP		1	11	100.00	>212.0	>212.0	10/30/2025	11:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6137713

WorkList Name : FLAHPPOINT

WorkList ID : 192765

Department : Wet-Chemistry

Date : 10-29-2025 16:30:17

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3479-01	92925	Solid	Flash Point	Cool 4 deg C	PSEG03	D41	10/28/2025	1010B

Date/Time 10/30/2025 09:00
 Raw Sample Received by: RM (wo)
 Raw Sample Relinquished by: JAC (wo)

Date/Time 10/30/2025 14:40
 Raw Sample Received by: JAC (wo)
 Raw Sample Relinquished by: RM (wo)