

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB137724

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/31/2025	08:50
2	CAL2	1	Water	NA	NA	20.2	7.00	10/31/2025	08:51
3	CAL3	1	Water	NA	NA	20.3	10.02	10/31/2025	08:52
4	ICV	1	Water	NA	NA	20.3	7.02	10/31/2025	08:55
5	CCV1	1	Water	NA	NA	20.3	2.01	10/31/2025	08:56
6	Q3496-03	1	Water	NA	NA	20.4	5.80	10/31/2025	09:10
7	Q3496-03DUP	1	Water	NA	NA	20.6	5.82	10/31/2025	09:14
8	Q3500-02	1	Water	NA	NA	23.2	8.40	10/31/2025	09:25
9	Q3500-03	1	Water	NA	NA	23.1	7.53	10/31/2025	09:33
10	Q3500-04	1	Water	NA	NA	20.6	7.80	10/31/2025	09:47
11	Q3500-05	1	Water	NA	NA	23.1	6.60	10/31/2025	09:50
12	Q3503-07	1	Water	NA	NA	22.1	8.27	10/31/2025	10:05
13	Q3505-04	1	Water	NA	NA	20.3	8.60	10/31/2025	10:15
14	Q3505-05	1	Water	NA	NA	20.1	10.84	10/31/2025	10:30
15	Q3505-06	1	Water	NA	NA	20.7	11.33	10/31/2025	10:40
16	CCV2	1	Water	NA	NA	20.2	12.02	10/31/2025	10:41

WORKLIST(Hardcopy Internal Chain)

UB 137724

WorkList Name : ph w q3492

WorkList ID : 192816

Department : Wet-Chemistry

Date : 10-31-2025 08:30:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3496-03	A PURGE-WATER	Water	pH	Cool 4 deg C	PSEG03	D41	10/30/2025	9040C
Q3500-02	D 462	Water	pH	Cool 4 deg C	PSEG03	D41	10/30/2025	9040C
Q3500-03	D 463	Water	pH	Cool 4 deg C	PSEG03	D41	10/30/2025	9040C
Q3500-04	A 465	Water	pH	Cool 4 deg C	PSEG03	D41	10/30/2025	9040C
Q3500-05	D 461-LIQUID	Water	pH	Cool 4 deg C	PSEG03	D41	10/30/2025	9040C
Q3503-07	A 320	Water	pH	Cool 4 deg C	PSEG03	D41	10/30/2025	9040C
Q3505-04	C 10-10-25-WATER	Water	pH	Cool 4 deg C	PSEG03	D41	10/30/2025	9040C
Q3505-05	C 10-16-25-WATER	Water	pH	Cool 4 deg C	PSEG03	D41	10/30/2025	9040C
Q3505-06	C 10-17-25-WATER	Water	pH	Cool 4 deg C	PSEG03	D41	10/30/2025	9040C

Date/Time 10/31/25 08:45

Raw Sample Received by: JB (WOC)

Raw Sample Relinquished by: RJ (FAA-Vet)

Date/Time 10/31/25 13:40

Raw Sample Received by: RJ (EXT-1066)

Raw Sample Relinquished by: JB (WOC)