

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137736

Slope : 98.2

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/03/2025	10:00
2	CAL2	1	Water	NA	NA	20.2	7.00	11/03/2025	10:01
3	CAL3	1	Water	NA	NA	20.3	10.02	11/03/2025	10:03
4	ICV	1	Water	NA	NA	20.3	7.02	11/03/2025	10:05
5	CCV1	1	Water	NA	NA	20.3	2.01	11/03/2025	10:10
6	Q3513-01	1	Solid	20.02	20	22.6	7.32	11/03/2025	10:20
7	Q3513-01DUP	1	Solid	20.03	20	22.7	7.34	11/03/2025	10:22
8	CCV2	1	Water	NA	NA	20.3	12.02	11/03/2025	10:25

WORKLIST(Hardcopy Internal Chain)

11/03/25

WorkList Name : PH S Q3513 WorkList ID : 192832 Department : Wet-Chemistry Date : 11-03-2025 07:42:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3513-01	A OILY-DEBRIS-DRUM	Solid	pH	Cool 4 deg C	PSEG03	D41	10/31/2025	9045D

Date/Time 11/03/25 07:50
Raw Sample Received by: SA WDC
Raw Sample Relinquished by: JT (SM)

Date/Time 11/03/25 13:00
Raw Sample Received by: JTCSA
Raw Sample Relinquished by: SA WDC