

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB137739

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

True Value of CCV3 = 2.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/03/2025	08:14
2	CAL2	1	Water	NA	NA	20.2	7.00	11/03/2025	08:15
3	CAL3	1	Water	NA	NA	20.3	10.02	11/03/2025	08:16
4	ICV	1	Water	NA	NA	20.3	7.01	11/03/2025	08:20
5	CCV1	1	Water	NA	NA	20.3	2.01	11/03/2025	08:25
6	Q3517-02	1	Water	NA	NA	22.3	7.62	11/03/2025	08:35
7	Q3517-02DUP	1	Water	NA	NA	22.5	7.64	11/03/2025	08:36
8	Q3517-03	1	Water	NA	NA	22.4	5.20	11/03/2025	08:40
9	Q3518-01	1	Water	NA	NA	22.6	7.71	11/03/2025	08:47
10	Q3518-02	1	Water	NA	NA	2.4	5.91	11/03/2025	08:55
11	Q3521-01	1	Water	NA	NA	22.6	5.28	11/03/2025	09:10
12	Q3521-02	1	Water	NA	NA	22.6	7.72	11/03/2025	09:20
13	Q3521-03	1	Water	NA	NA	22.6	7.52	11/03/2025	09:22
14	Q3524-01	1	Water	NA	NA	22.6	4.75	11/03/2025	09:37
15	Q3524-02	1	Water	NA	NA	22.6	5.03	11/03/2025	09:45
16	CCV2	1	Water	NA	NA	20.2	12.02	11/03/2025	09:47
17	Q3524-03	1	Water	NA	NA	22.7	5.05	11/03/2025	09:55
18	Q3524-04	1	Water	NA	NA	22.6	5.09	11/03/2025	10:00
19	CCV3	1	Water	NA	NA	20.3	2.02	11/03/2025	10:05

WORKLIST(Hardcopy Internal Chain)

137739

WorkList Name : PH W Q3524 WorkList ID : 192831 Department : Wet-Chemistry Date : 11-03-2025 07:41:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3517-02	A NWB-2156	Water	pH	Cool 4 deg C	PSEG03	D41	10/31/2025	9040C
Q3517-03	A NWB-2226	Water	pH	Cool 4 deg C	PSEG03	D41	10/31/2025	9040C
Q3518-01	A 50771	Water	pH	Cool 4 deg C	PSEG03	D41	10/31/2025	9040C
Q3518-02	A 50831	Water	pH	Cool 4 deg C	PSEG03	D41	10/31/2025	9040C
Q3521-01	C 102325	Water	pH	Cool 4 deg C	PSEG03	D41	10/31/2025	9040C
Q3521-02	A 103125	Water	pH	Cool 4 deg C	PSEG03	D41	10/31/2025	9040C
Q3521-03	B TANK-2025	Water	pH	Cool 4 deg C	PSEG03	D41	10/31/2025	9040C
Q3524-01	B TH-430-DI-1	Water	pH	Cool 4 deg C	NYUL01	D41	10/31/2025	9040C
Q3524-02	B TH-430-DI-2	Water	pH	Cool 4 deg C	NYUL01	D41	10/31/2025	9040C
Q3524-03	B TH-430-DI-3	Water	pH	Cool 4 deg C	NYUL01	D41	10/31/2025	9040C
Q3524-04	B TH-404-DI-4	Water	pH	Cool 4 deg C	NYUL01	D41	10/31/2025	9040C

Date/Time 11/03/25 08:00
 Raw Sample Received by: 88 WCC
 Raw Sample Relinquished by: JTC547

Date/Time 11/03/25 13:30
 Raw Sample Received by: JTC547
 Raw Sample Relinquished by: 88 WCC